

VIASURE

Real Time PCR Detection Kits

by CerTest
BIOTEC

Human Herpes Virus 6, 7 & 8

Handbook for the following references/
Manual para las siguientes referencias:

VIASURE Human Herpes Virus 6, 7 & 8 Real Time PCR Detection Kit 6 x 8-well strips, low profile	VS-HHV106L
VIASURE Human Herpes Virus 6, 7 & 8 Real Time PCR Detection Kit 6 x 8-well strips, high profile	VS-HHV106H
VIASURE Human Herpes Virus 6, 7 & 8 Real Time PCR Detection Kit 12 x 8-well strips, low profile	VS-HHV112L
VIASURE Human Herpes Virus 6, 7 & 8 Real Time PCR Detection Kit 12 x 8-well strips, high profile	VS-HHV112H
VIASURE Human Herpes Virus 6, 7 & 8 Real Time PCR Detection Kit 96-well plate, low profile	VS-HHV113L
VIASURE Human Herpes Virus 6, 7 & 8 Real Time PCR Detection Kit 96-well plate, high profile	VS-HHV113H
VIASURE Human Herpes Virus 6, 7 & 8 Real Time PCR Detection Kit 9 x 4-well strips, Rotor-Gene®	VS-HHV136
VIASURE Human Herpes Virus 6, 7 & 8 Real Time PCR Detection Kit 18 x 4-well strips, Rotor-Gene®	VS-HHV172



ENGLISH

1. Intended use

VIASURE Human Herpes Virus 6, 7 & 8 Real Time PCR Detection Kit is designed for the specific identification and differentiation of Human Herpesvirus 6, Human Herpesvirus 7 and/or Human Herpesvirus 8 in clinical samples from patients with signs and symptoms of Human Herpesvirus 6, Human Herpesvirus 7 and Human Herpesvirus 8 infection. This test is intended to be used as an aid in the diagnosis of HHV-6, HHV-7 and/or HHV-8 in combination with clinical and epidemiological risk factors. DNA is extracted from specimens, multiplied using Real Time amplification and detected using fluorescent reporter dye probes specific for HHV-6, HHV-7 and HHV-8.

2. Summary and Explanation

Herpesviridae is a large family of DNA viruses that cause diseases in humans. The members of this family are also known as herpesviruses. There are 9 herpesvirus types known to infect humans: herpes simplex viruses 1 and 2, HSV-1 and HSV-2, (HHV1 and HHV2), varicella-zoster virus (VZV or HHV-3), Epstein-Barr virus (EBV or HHV-4), human cytomegalovirus (HCMV or HHV-5), human herpesvirus 6A and 6B (HHV-6A and HHV-6B), human herpesvirus 7 (HHV-7), and Kaposi's sarcoma-associated herpesvirus (KSHV or HHV-8).

Human herpesviruses 6 (HHV-6) A and B are two distinct lymphotropic β -herpesviruses that infect the majority of humans. By age 5, the majority of humans have acquired HHV-6 infection as indicated by seropositivity. The classic presentation of primary HHV-6 infection is Roseola Infantum. In immunocompetent individuals, primary HHV-6 infection is often asymptomatic or it could be manifested clinically as a febrile illness associated with an exanthema, diarrhea, respiratory symptoms, seizures or encephalitis.

Human herpesvirus (HHV-7) is also a lymphotropic β -herpesvirus that infects humans during the first 5 years of life. Like HHV-6, it may be asymptomatic or it could manifest clinically as fever, upper respiratory symptoms, vomiting, diarrhea, seizures, encephalitis and a Roseola-like illness. Even though, the most common clinical presentations in HHV-7 viremic children are seizures and fever.

Latency of β -herpesviruses is established in secretory glands, lymphoreticular cells, and in tissues such as the kidneys among others. Plasma, serum, whole blood, peripheral blood mononuclear cell (PBMC) preparations, saliva and cerebrospinal fluid (CSF) specimens from infants and immunocompromised patients are acceptable specimen types for HHV-6 and HHV-7 PCR based detection and monitoring.

Human herpesvirus 8 (HHV-8) is a γ -herpesvirus Kaposi's sarcoma-associated herpesvirus. It is transmitted during intimate contact, likely through transfer of body secretions such as saliva and genital secretions. HHV-8 can also be transmitted through blood transfusion and organ transplantation. In immunocompetent children, primary HHV-8 infection is associated with mild nonspecific symptoms of fever and rash. In adults, primary HHV-8 infection has been implicated in clinical illness characterized by diarrhea, fatigue and lymphadenopathy. Following resolution of primary infection, HHV-8 enters a state of latency in cells of lymphoid origin. Biopsies of lesion (oral and cutaneous), saliva, blood, plasma, and PBMC are the clinical specimens more frequently used for viral detection and monitoring of treatment of HHV-8.

Infections due to HHV-6 and HHV-7 are common in transplant recipients (i.e. solid organ transplant (SOT)), but clinical diseases due to these viruses are uncommon. Fever, rash and bone marrow suppression are the most common manifestations, but symptoms of tissue invasive disease may be observed. Infections due to HHV-8 may



lead to Kaposi's sarcoma and less commonly Castleman's disease, primary effusion lymphomas, bone marrow suppression and multiorgan failure. HHV-8 disease is most common among transplant recipients from endemic regions, although certain populations in nonendemic regions are also at high risk.

The amplification of the viral DNA by the polymerase chain reaction is used extensively throughout the world for its sensitivity and specificity. qPCR is an ideal technique to monitor current infections.

3. Principle of the procedure

VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit is designed for the diagnosis of HHV-6, HHV-7 and HHV-8 in clinical samples. After DNA isolation, the identification of Human Herpesvirus 6, Human Herpesvirus 7 and Human Herpesvirus 8 is performed by the amplification of a conserved region of the *U65-U66* genes for Human Herpesvirus 6, *ORF73* gene for Human Herpesvirus 8, and *U57* gene for Human Herpesvirus 7, using specific primers and a fluorescent-labelled probe.

VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit is based on the 5' exonuclease activity of DNA polymerase. During DNA amplification, this enzyme cleaves the probe bounded to the complementary DNA sequence, separating the quencher dye from the reporter. This reaction generates an increase in the fluorescent signal which is proportional to the quantity of target template. This fluorescence can be measured on Real Time PCR platforms.

VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit contains in each well all the components necessary for real time PCR assay (specific primers/probes, dNTPS, buffer, polymerase) in an stabilized format, as well as an internal control to monitor PCR inhibition. Human Herpesvirus 6 DNA targets are amplified and detected in the FAM channel, Human Herpesvirus 8 DNA targets are amplified and detected in the ROX channel, Human Herpesvirus 7 DNA targets are amplified and detected in the Cy5 channel and the internal control (IC) in HEX, VIC or JOE channel (depending on the equipment used select the proper detection channel, see Annex 2).

4. Reagents provided

VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit includes the following materials and reagents detailed in Tables 1, 2 and 3. Based on the commercial presentation and the Real Time PCR platform used, the stabilized PCR reaction mix could be placed inside different wells and could be marketed on multiple formats. Table 1 includes materials and reagents to be used with 8-well strips compatible devices (See Annex 1). Table 2 includes materials and reagents to be used with 96-well plate compatible devices (See Annex 1). Table 3 includes materials and reagents for use with Qiagen/Corbett Rotor-Gene® instruments for 4-well strips.



Reagent/Material	Description	Colour	Amount
Human Herpes Virus 6, 7 & 8 8-well strips	A mix of enzymes, primers probes, buffer, dNTPs, stabilizers and Internal control in stabilized format	White	6/12 x 8-well strip
Rehydration Buffer	Solution to reconstitute the stabilized product	Blue	1 vial x 1.8 mL
Human Herpes Virus 6, 7 & 8 Positive Control	Non-infectious synthetic lyophilized DNA	Red	1 vial
Negative control	Non template control	Violet	1 vial x 1 mL
Water RNase/DNase free	RNase/DNase free water	White	1 vial x 1 mL
Tear-off 8-cap strips	Optical caps for sealing wells during thermal cycling	Transparent	6/12 X 8-cap strip

Table 1. Reagents and materials provided in VIASURE Human Herpes Virus 6, 7 & 8 Real Time PCR Detection Kit with Ref. VS-HHV106L, VS-HHV106H, VS-HHV112L and VS-HHV112H.

Reagent/Material	Description	Color	Amount
Human Herpes Virus 6, 7 & 8 96-well plate	A mix of enzymes, primers probes, buffer, dNTPs, stabilizers and Internal control in stabilized format	White	1 plate
Rehydration Buffer	Solution to reconstitute the stabilized product	Blue	1 vial x 1.8 mL
Human Herpes Virus 6, 7 & 8 Positive Control	Non-infectious synthetic lyophilized DNA	Red	1 vial
Negative control	Non template control	Violet	1 vial x 1 mL
Water RNase/DNase free	RNase/DNase free water	White	1 vial x 1 mL
Tear-off 8-cap strips	Optical caps for sealing plate during thermal cycling	Transparent	12 X 8-cap strip

Table 2. Reagents and materials provided in VIASURE Human Herpes Virus 6, 7 & 8 Real Time PCR Detection Kit with Ref VS-HHV113L and VS-HHV113H.

Reagent/Material	Description	Colour	Amount
Human Herpes Virus 6, 7 & 8 4-well strips	A mix of enzymes, primers probes, buffer, dNTPs, stabilizers and Internal control in stabilized format	Transparent	9/18 x 4-well strip
Rehydration Buffer	Solution to reconstitute the stabilized product	Blue	1 vial x 1.8 mL
Human Herpes Virus 6, 7 & 8 Positive Control	Non-infectious synthetic lyophilized DNA	Red	1 vial
Negative control	Non template control	Violet	1 vial x 1 mL
Water RNase/DNase free	RNase/DNase free water	White	1 vial x 1 mL
4-cap strips	Optical caps for sealing wells during thermal cycling	Transparent	9/18 X 4-cap strip

Table 3. Reagents and materials provided in VIASURE Human Herpes Virus 6, 7 & 8 Real Time PCR Detection Kit with Ref. VS-HHV136 and VS-HHV172. For use with Qiagen/Corbett Rotor-Gene® instruments and compatible accessories with strips of 4 tubes 0.1 ml (72-Well Rotor and Locking Ring 72-Well Rotor).

5. Reagents and equipment to be supplied by the user

The following list includes the materials that are required for use but not included in the VIASURE Human Herpes Virus 6, 7 & 8 Real Time PCR Detection Kit.



- Real Time PCR instrument (thermocycler).
- DNA extraction kit.
- Centrifuge for 1.5 mL tubes and PCR-well strips or 96-well plate (if available).
- Vortexer.
- Micropipettes (0.5-20 µL, 20-200 µL).
- Filter tips.
- Powder-free disposable gloves.
- Loading block (for use with Qiagen/Corbett Rotor-Gene® instruments).

VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit has been validated on the following equipments: Applied Biosystems 7500 Fast Real-Time PCR System, Bio-Rad CFX96™ Real-Time PCR Detection System, Agilent Technologies AriaMx Real-Time PCR System, DNA-Technology DTprime Real-time Detection Thermal Cycler, DNA-Technology DTlite Real-time Detection Thermal Cycler, Rotor-Gene® Q (Qiagen), SmartCycler® (Cepheid), Roche Molecular Diagnostics Cobas z480 Analyzer, VIASURE 48 Real Time PCR System and VIASURE 96 Real Time PCR System. When using the Applied Biosystems 7500 Fast with strips it is recommend to place a plate holder to reduce the risk of crushed tube (Ref. PN 4388506).

To check thermocycler compatibility, see Annex 1, to check most common detection channels see Annex 2 and to check optical measurement exposure setting see Annex 3.

6. Transport and storage conditions

- The kits can be shipped and stored at 2-40°C until the expiration date which is stated on the label.
- Once the positive control has been re-suspended, store it at -20°C. We recommend to separate it in aliquots to minimize freeze and thaw cycles. Positive control has been validated as still being stable after 6 freeze-thaw cycles.
- Keep components away from sunlight.

7. Precautions for users

- The product is indented for use by professional users only, such as laboratory or health professionals and technicians, trained in molecular biological techniques.
- Do not use past expiration date.
- Do not use reagents if the protective pouches are open or broken upon arrival.
- Do not use reagents if desiccant is not present or broken inside reagent pouches.
- Do not remove desiccant from reagent pouches once is open.
- Close protective pouches of reagents promptly with the zip seal after each use (if available, Ref. VS-HHV113L, VS-HHV113H, VS-HHV136 and VS-HHV172). Remove any excess air in the pouches prior to sealing.
- Do not use reagents if the foil has been broken or damaged.
- Do not mix reagents from different envelopes and / or kits and / or lots and / or another supplier.
- Protect reagents against from humidity. Prolonged exposure to humidity may affect product performance.



- For references VS-HHV136 and VS-HHV172 (compatible with Qiagen/Corbett Rotor-Gene® instruments) use the loading block to pipette reagents and samples into each tube and to help with fitting caps properly and avoid cross contamination.
- Design a unidirectional workflow. It should begin in the Extraction Area and then move to the Amplification and Detection Area. Do not return samples, equipment and reagents to the area in which the previous step was performed.
- Follow Good Laboratory Practices. Wear protective clothing, use disposable gloves, goggles and mask. Do not eat, drink or smoke in the working area. Once you finish the test wash your hands.
- Specimens must be treated as potentially infectious, as well as all the reagents and materials that have been exposed to the samples and they must be handled according to the national safety regulations. Take necessary precautions during the collection, storage, treatment and disposal of samples.
- Regular decontamination of commonly used equipment is recommended, especially micropipettes and work surfaces.
- Consult safety data sheets, upon request.
- Consult each Real Time PCR instrument's reference manual for additional warnings, precautions and procedures.

8. Test procedure

8.1. DNA extraction

Perform the sample preparation according to the recommendations appearing in the instructions for use of the extraction kit used.

For DNA extraction from clinical samples you can use your manual or automatic routine optimized system. Also, you can use any commercially available DNA extraction kit and follow the manufacturer's instructions. We have validated the following extraction kits:

- Viasure RNA-DNA Extraction kit (VIASURE), recommended.
- Invisorb® Spin Universal Kit (Strattec).

8.2. Lyophilized positive control

Human Herpes Virus 6, 7 & 8 Positive Control contains high copies of the template, the recommendation is to open and manipulate it in a separate laboratory area away from the other components. Reconstitute the lyophilized *Human Herpes Virus 6, 7 & 8* Positive Control (red vial) by adding 100 µL of the supplied Water RNase/DNase free (white vial) and vortex thoroughly.

Once the positive control has been re-suspended, store it at -20°C. We recommend to separate it in aliquots to minimize freeze and thaw cycles.



8.3. PCR protocol

Determine and separate the number of required reactions including samples and controls. One positive and negative control must be included in each run for each assay. Peel off protective aluminium seal from plates or strips.

- 1) Reconstitute the number of wells you need.

Add 15 µL of Rehydration Buffer (blue vial) into each well.

- 2) Adding samples and controls.

Add 5 µL of DNA sample, reconstituted *Human Herpes Virus 6, 7 & 8* Positive Control (red vial) or Negative Control (violet vial) in different wells and close them with the provided caps.

It is recommended to briefly centrifuge the 8-well strips or 96-well plate, or gently tap each strip onto a hard surface to ensure that all the liquids are at the bottom of the tubes (for Qiagen/Corbett Rotor-Gene®).

Load the plate or the strips in the thermocycler.

- 3) Set up the thermocycler (to check compatibility see Annex 1).

Program the thermocycler following the conditions listed below and start the run:

Cycles	Step	Time	Temperature
1	Initial denaturation	2 min	95°C
45	Denaturation	10 seg	95°C
	Annealing/Extension (Data collection*)	50 seg	60°C

Table 4. PCR protocol

Fluorogenic data should be collected during the extension step (*) through the FAM (Human Herpesvirus 6), ROX (Human Herpesvirus 8), Cy5 (Human Herpesvirus 7) and HEX, JOE or VIC channels (CI). Depending on the equipment used select the proper detection channel (see Annex 2). In the Applied Biosystems 7500 Fast Real-Time PCR System and Stratagene Mx3005P™ Real Time PCR System check that the passive reference option for ROX is none. In the Applied Biosystems 7500 Fast Real-Time PCR System select Ramp Speed Standard in Select New Experiment/Advanced Setup/Experiment Properties.

9. Result interpretation

The use of positive and negative controls in each run, validate the reaction by checking the absence of signal in the negative control well and the presence of signal for *Human Herpes Virus 6, 7 & 8* in the positive control well. Check Internal Control signal to verify the correct functioning of the amplification mix. The analysis of the samples is done by the software of the used real time PCR equipment itself according to manufacturer's instructions.

Using the following table read and analyze the results:



Human Herpesvirus 6 (FAM)	Human Herpesvirus 8 (ROX)	Human Herpesvirus 7 (Cy5)	Internal control (HEX)	Negative Control	Positive Control	Interpretation
+	+	+	+/-	-	+	Human Herpesvirus 6, Human Herpesvirus 8 and Human Herpesvirus 7 Positive
-	-	-	+	-	+	Human Herpesvirus 6, Human Herpesvirus 8 and Human Herpesvirus 7 Negative
+	-	-	+/-	-	+	Human Herpesvirus 6 Positive, Human Herpesvirus 8 and Human Herpesvirus 7 Negative
+	+	-	+/-	-	+	Human Herpesvirus 6 and Human Herpesvirus 8 Positive, and Human Herpesvirus 7 Negative
+	-	+	+/-	-	+	Human Herpesvirus 6 and Human Herpesvirus 7 Positive, and Human Herpesvirus 8 Negative
-	+	-	+/-	-	+	Human Herpesvirus 8 Positive, Human Herpesvirus 6 and Human Herpesvirus 7 Negative
-	+	+	+/-	-	+	Human Herpesvirus 8 and Human Herpesvirus 7 Positive, Human Herpesvirus 6 Negative
-	-	+	+/-	-	+	Human Herpesvirus 7 Positive, Human Herpesvirus 6 and Human Herpesvirus 8 Negative
-	-	-	-	-	+	Experiment fail
+	+	+	+	+	-	Experiment fail

Table 5. Sample interpretation

+: Amplification curve


-: No amplification curve

A sample is considered positive if the Ct value obtained is less than 40 and the internal control shows or not an amplification signal. Sometimes, the detection of internal control is not necessary because a high copy number of target can cause preferential amplification of target-specific nucleic acids

A sample is considered negative, if the sample shows no amplification signal in the detection system but the internal control is positive. An inhibition of the PCR reaction can be excluded by the amplification of internal control.



Figure 1. Correct run of negative and positive control run on the Bio-Rad CFX96™ Real-Time PCR Detection System.



The result is considered invalid if there is signal of amplification in negative control or absence of signal in the positive well. We recommend to repeat the assay again.

In case of absence of internal control signal in sample wells we recommend to repeat the assay diluting the sample 1:10 or to repeat the extraction to check for possible problems of inhibition.

In case of a doubtful interpretation result, it is recommended to verify the correct performance of each of the steps and review the parameters and the sigmoid shape of the curve. If the situation is not solved, it is recommended to repeat the assay, preferably in duplicate. The results of the test should be evaluated by a health care professional in the context of medical history, clinical symptoms and other diagnostic tests.

10. Limitations of the test

- The results of the test should be evaluated by a health care professional in the context of medical history, clinical symptoms and other diagnostic tests.
- Although this assay can be used with other types of samples, it has been validated only with culture, transport medium, plasma and CFS samples.
- The quality of the test depends on the quality of the sample; proper extracted nucleic acid from clinical samples must be extracted. Unsuitable collection, storage and/or transport of specimens may give false negative results.
- Extremely low levels of target below the limit of detection might be detected, but results may not be reproducible.
- There is a possibility of false positive results due to cross-contamination by Human Herpesvirus 6, Human Herpesvirus 7 and Human Herpesvirus 8 either by samples containing high concentrations of target DNA or contamination due to PCR products from previous reactions.



11. Quality control

VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit contains a positive and a negative control that must be included in each run to correctly interpret the results. Also, the internal control (IC) in each well confirms the correct performance of the technique.

12. Performance characteristics

12.1. Clinical sensitivity and specificity

The clinical performance of VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit was evaluated using different EQA panels from QCMD. These panels consist of 20 frozen plasma samples or transport medium samples. The results were compared with the QCMD final reports. 9 of 10 samples from the Human Herpes Virus 6 panel were positive for the pathogen and 1 of 10 samples from the Transplantation Panel were positive for HHV-6 using VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit.

In conclusion, the results show a high sensitivity and specificity to detect Human Herpesvirus 6 using VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit.

12.2. Analytical sensitivity

VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit has a detection limit of ≥ 10 DNA copies per reaction for Human Herpesvirus 6, Human Herpesvirus 7 and Human Herpesvirus 8 (Figure 2, 3 and 4).

Figure 2. Dilution series of Human Herpes Virus 6 (10^7 - 10^1 copies/rxn) template run on the Bio-Rad CFX96™ Real-Time PCR Detection System (channel FAM).




Figure 3. Dilution series of Human Herpes Virus 7 (10^7 - 10^1 copies/rxn) template run on the Bio-Rad CFX96™ Real-Time PCR Detection System (channel Cy5).





Figure 4. Dilution series of Human Herpes Virus 8 (10^7 - 10^1 copies/rxn) template run on the Bio-Rad CFX96™ Real-Time PCR Detection System (channel ROX).



12.3. Analytical specificity

The specificity of the Human Herpesvirus 6, Human Herpesvirus 7 and Human Herpesvirus 8 assay was confirmed by testing a panel consisting of different microorganisms representing the most common pathogens that can cause meningitis and affect immunosuppressive patients in particular. No cross-reactivity was detected between almost any of the following microorganisms tested, except the targeted pathogens of each assay.



Cross-reactivity testing					
<i>E. coli</i> 0.1285;O18:H7:K1	-	<i>Streptococcus pneumoniae</i> Z022	-	St Louis Encephalitis Virus strain 17D	-
<i>Listeria innocua</i> Serotype 6a/strain CCUG 15531	-	<i>Plasmodium falciparum</i> 3D7	-	Tick-Borne encephalitis virus strain Neudorfl	-
<i>Listeria ivanovii</i> Serovar 5/strain CCUG 15528	-	<i>Cryptococcus gatti</i> Z156	-	Chikungunya virus (S27 Petersfield)	-
<i>Listeria monocytogenes</i> Serotype 1/2b	-	<i>Haemophilus influenzae</i> MinnA	-	Dengue 1 (Hawaii A strain)	-
<i>Listeria monocytogenes</i> Serovar 4b/Strain CIP 59.53	-	<i>Neisseria meningitidis</i> serogroup A	-	Dengue 2 (New Guinea C strain)	-
<i>Bacteroides fragilis</i>	-	<i>Treponema pallidum</i>	-	Dengue 3 (H87)	-
<i>Pseudomonas aeruginosa</i>	-	<i>Toxoplasma gondii</i> Type II	-	Dengue 4 (H241)	-
<i>Yersinia enterocolitica</i> O:3	-	Enterovirus type 68	-	West Nile Virus (Heja)	-
<i>Klebsiella oxytoca</i>	-	Enterovirus type 71	-	West Nile Virus (NY99)	-
<i>Enterobacter cloacae</i> Serotype Cloaca A	-	Epstein-Barr virus	-	West Nile Virus (Ug37)	-
<i>Enterobacter aerogenes</i> Serotype Cloaca B	-	BK Virus Type Ib2	-	Yellow Fever Virus (17D strain)	-
<i>Enterococcus durans</i>	-	BK Virus Type IV	-	Echovirus Type 30	-
<i>Enterococcus faecalis</i>	-	JC Virus Type 2B	-	Echovirus Type 11	-
<i>Enterococcus faecium</i> serotype 11	-	JC Virus Type 1A	-	HSV-1 strain MacIntyre	-
<i>Acinetobacter baumannii</i>	-	Citomegalovirus strain AD-169	-	HSV-2 MS	-
<i>Candida albicans</i>	-	Coxsackievirus type A24	-	HHV6 strain Z29	-/+
<i>Aspergillus fumigatus</i>	-	Coxsackievirus type A9	-	HHV6 Type A	-/+
<i>Pneumocystis jirovecii</i> Type A1	-	Coxsackievirus type B3	-	HHV6 Type B	-/+
<i>Pneumocystis jirovecii</i> g885652	-	Parechovirus Type 3	-	Varicella-Zoster Virus Ellen	-
<i>Staphylococcus aureus</i> subsp <i>aureus</i>	-	Parvovirus B19	-	Varicella Zoster Virus OKA	-
<i>Streptococcus agalactiae</i> Z019	-				

Table 6. Reference pathogenic microorganisms used in this study.

12.4. Analytical reactivity

The reactivity of VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit for Human Herpesvirus 6 was evaluated against HHV6 strain Z29, HHV6 Type A and HHV6 Type B showing positive result.

The reactivity of VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit for Human Herpesvirus 8 was evaluated against Human Herpesvirus 8 synthetic sequence, showing positive result.

The reactivity of VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit for Human Herpesvirus 7 was evaluated against Human Herpesvirus 7 synthetic sequence, showing positive result.



ANNEX 1

COMPATIBILITY WITH THE MOST COMMON REAL TIME PCR EQUIPMENT

Low profile strips can be used in all PCR thermocyclers equipped with a low profile block, like the systems listed in table A.1. High profile strips can be used in all PCR thermocyclers equipped with a high or regular profile block, like the systems listed in table A.2. If you do not find your thermocycler in the list below, please contact with your supplier.

Table A.1 LOW PROFILE BLOCK THERMOCYCLERS	
Manufacturer	Model
Agilent Technologies	AriaMx/AriaDx Real-Time PCR System
Applied Biosystems	7500 Fast Real-Time PCR System ⁽¹⁾
Applied Biosystems	7500 Fast Dx Real-Time PCR System ⁽¹⁾
Applied Biosystems	QuantStudio™ 12K Flex 96-well Fast
Applied Biosystems	QuantStudio™ 6 Flex 96-well Fast
Applied Biosystems	QuantStudio™ 7 Flex 96-well Fast
Applied Biosystems	QuantStudio™ 3 Fast Real-Time PCR System ⁽⁵⁾
Applied Biosystems	QuantStudio™ 5 Fast Real-Time PCR System
Applied Biosystems	StepOne Plus™ Real-Time PCR System ⁽⁵⁾
Applied Biosystems	StepOne™ Real-Time PCR System ⁽⁵⁾
Applied Biosystems	ViiA™ 7 Fast Real-Time PCR System
BIONEER	Exicycler™ 96
Bio-Rad	CFX96™ / CFX96™ IVD Real-Time PCR Detection System
Bio-Rad	Mini Opticon™ Real-Time PCR Detection System ⁽⁶⁾
Cepheid	SmartCycler® ⁽³⁾
Qiagen	Rotor-Gene® Q ⁽³⁾
Roche	LightCycler® 480 Real-Time PCR System ⁽⁴⁾
Roche	LightCycler® 96 Real-Time PCR System ⁽⁴⁾
Roche	Cobas z480 Analyzer ⁽⁴⁾

Table A.2 HIGH PROFILE BLOCK THERMOCYCLERS	
Manufacturer	Model
Abbott	Abbott m2000 RealTime System
Applied Biosystems	7300 Real-Time PCR System ⁽⁵⁾
Applied Biosystems	7500 Real-Time PCR System
Applied Biosystems	7900 HT Real-Time PCR System ⁽⁵⁾
Applied Biosystems	ABI PRISM 7000 ⁽⁶⁾
Applied Biosystems	ABI PRISM 7700 ⁽⁵⁾
Applied Biosystems	QuantStudio™ 12K Flex 96-well
Applied Biosystems	QuantStudio™ 6 Flex 96-well
Applied Biosystems	QuantStudio™ 7 Flex 96-well
Applied Biosystems	QuantStudio™ 3 Real-Time PCR System ⁽⁵⁾
Applied Biosystems	QuantStudio™ 5 Real-Time PCR System
Applied Biosystems	ViiA™ 7 Real-Time PCR System
Analytik Jena Biometra	TOptical
Analytik Jena Biometra	qTOWER 2.0
BIONEER	Exicycler™ 96
Bio-Rad	CFX96™ Deep Well / CFX96™ Deep Well IVD Real-Time PCR Detection System
Bio-Rad	iCycler iQ™ Real-Time PCR Detection System
Bio-Rad	iCycler iQ™5 Real-Time PCR Detection System
Bio-Rad	MyiQ™ Real-Time PCR Detection System ⁽⁶⁾
Bio-Rad	MyiQ™2 Real-Time PCR Detection System ⁽⁶⁾
Cepheid	SmartCycler® ⁽³⁾
DNA-Technology	DTprime Real-time Detection Thermal Cycler ⁽²⁾
DNA-Technology	DTlite Real-Time PCR System ⁽²⁾
Eppendorf	Mastercycler™ep realplex
Qiagen	Rotor-Gene® Q ⁽³⁾
Stratagene / Agilent Technologies	Mx3000P™ Real Time PCR System
Stratagene / Agilent Technologies	Mx3005P™ Real Time PCR System
VIASURE	VIASURE 48 Real Time PCR System ⁽²⁾
VIASURE	VIASURE 96 Real Time PCR System ⁽²⁾

(1) Select Ramp Speed "Standard".

(2) See Annex 3 to check optical measurement exposure setting.

(3) The product should be reconstituted following the appropriate procedure (see Test Procedure) and transferred into the specific Rotor-Gene® Q or SmartCycler® tubes.

(4) Shell Frame grid plate which fits in these Roche qPCR System is necessary.

(5) No detection in Cy5 channel.

(6) Detection in FAM and HEX channels only.

Table A1/A2. Compatible low and high profile Real Time PCR systems.



ANNEX 2

DETECTION CHANNELS FOR THE MOST COMMON REAL TIME PCR EQUIPMENT

The fluorescence detection channels for some of most common Real Time PCR Thermocyclers are specified in Table A3.

REAL-TIME PCR THERMOCYCLER	VIASURE CHANNEL	DETECTION CHANNEL	OBSERVATIONS
Bio-Rad CFX96™	FAM	FAM	Some wells may have abnormally drifting RFU values during the initial few cycles of a run showing a non-sigmoidal ascendant line. If you see this effect, in the Settings menu, select the option Apply Fluorescence Drift Correction for Baseline Settings to correct it.
	HEX	HEX	
	ROX	ROX	
	Cy5	Cy5	
ABI 7500 Applied Biosystems	FAM	FAM	Passive reference option for ROX must be none. Some wells may have abnormally drifting RFU values during the initial few cycles of a run showing a non-sigmoidal ascendant line. If you see this effect, please modify the baseline: Select the Start Cycle and End Cycle values so that the baseline ends before significant fluorescence is detected.
	HEX	VIC	
	ROX	ROX	
	Cy5	Cy5	
Roche Lightcycler®480II	FAM	465/510	Colour Compensation is required
	HEX	533/580	
	ROX	533/610	
	Cy5	618/660	
Smartcycler® Cepheid	FAM	Channel 1	
	HEX	Channel 2	
	ROX	Channel 3	
	Cy5	Channel 4	
Abbott m2000rt	FAM	FAM	
	HEX	VIC	
	ROX	ROX	
	Cy5	Cy5	
Mx3000P™ Mx 3005P™ Stratagene	FAM	FAM	Passive reference option for ROX must be none
	HEX	VIC	
	ROX	ROX	
	Cy5	Cy5	
AriaMx Agilent	FAM	FAM	
	HEX	HEX	
	ROX	ROX	
	Cy5	Cy5	
Rotor-Gene®Q Qiagen	FAM	Green	In the Channel Setup, click on the "Gain Optimisation" button and then go to "Optimise Acquiring". The fluorescence Target Sample Range has to be between 5 and 10 FI for each channel. Also select the option "Perform Optimisation Before 1st Acquisition".
	HEX	Yellow	
	ROX	Orange	
	Cy5	Red	
Mic Real Time PCR Cycler bms	FAM	Green	In the "Run Profile" menu, introduce the correct parameters for "Temperature Control" (Standard TAQ (v3)), Volume (20 ul) and the appropriate thermal profile. In the "Cycling" window, select the "Acquire on" option for all the channels by clicking on them. Use the default "Gain" values for each channel (Green = 3, Yellow = 10, Orange = 10, Red = 10)
	HEX	Yellow	
	ROX	Orange	
	Cy5	Red	
Exicycler™ 96 BIONEER	FAM	FAM	
	HEX	JOE	
	ROX	ROX	
	Cy5	Cy5	

Table A3: Detection fluorescence channels of different Real Time PCR systems.



ANNEX 3

OPTICAL MEASUREMENT EXPOSURE SETTING

Optical measurement parameters of some thermocyclers must be adjusted to be suitable for operation with "VIASURE Real Time PCR Detection Kits". This assay has been validated with the following set exposition values:

- DTprime Real-time Detection Thermal Cycler (DNA-Technology) and VIASURE 96 Real Time PCR System (CerTest Biotec S.L.): FAM channel -500*, HEX channel – 1000, ROX channel – 1000 and Cy5 channel - 1000.
- DTlite Real-Time PCR System (DNA-Technology) and VIASURE 48 Real Time PCR System (CerTest Biotec S.L.): FAM channel - 500, HEX channel - 500, ROX channel – 500 and Cy5 channel - 500.

*If the result in channel FAM is not as expected, there are no amplifications or high background noise is observed, please lower the exposure values indicated above to 150.



ESPAÑOL

1. Uso previsto

VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit está diseñado para la identificación y diferenciación específica del Herpesvirus humano 6, Herpesvirus humano 7 y Herpesvirus humano 8 en muestras clínicas procedentes de pacientes con signos y síntomas de infección por los Herpesvirus humano 6, Herpesvirus humano 7 y Herpesvirus humano 8. El uso previsto del test es facilitar el diagnóstico de infección producida por los HHV-6, HHV-7 y/o HHV-8 en combinación con factores de riesgos clínicos y epidemiológicos. El DNA es extraído a partir de las muestras, amplificado posteriormente mediante PCR a tiempo real y detectado utilizando oligonucleótidos específicos y una sonda marcada con una molécula fluorescente y otra apantalladora (*quencher*) para detectar HHV-6, HHV-7 y HHV-8.

2. Introducción y explicación

Herpesviridae es una gran familia de virus de DNA que causan enfermedades en humanos. Los miembros de esta familia también se conocen como herpesvirus. Hay 9 tipos de virus del herpes que se sabe que infectan a los humanos: virus del herpes simple 1 y 2, HSV-1 y HSV-2, (HHV-1 y HHV-2), virus varicela-zoster (VZV o HHV-3), virus de Epstein-Barr (EBV o HHV-4), citomegalovirus humano (HCMV o HHV-5), herpesvirus humano 6A y 6B (HHV-6A y HHV-6B), herpesvirus humano 7 (HHV-7) y herpesvirus asociado con sarcoma de Kaposi (KSHV o HHV-8).

Los herpesvirus humanos 6 (HHV-6) A y B son dos β -herpesvirus linfótrofos que infectan a la mayoría de los humanos. A la edad de 5 años, la mayoría de los seres humanos han adquirido la infección HHV-6, resultando seropositivos a dicho virus. La presentación clásica de la infección primaria HHV-6 es roséola infantum. En individuos inmunocompetentes, la infección primaria por HHV-6 a menudo es asintomática o puede manifestarse clínicamente como una enfermedad febril asociada con un exantema, diarrea, síntomas respiratorios, convulsiones o encefalitis.

El virus del herpes humano 7 (HHV-7) también es un β -herpesvirus linfotrópico que infecta a los seres humanos durante los primeros 5 años de vida. Al igual que HHV-6, su infección puede resultar asintomática o manifestarse clínicamente en forma de fiebre, síntomas respiratorios superiores, vómitos, diarrea, convulsiones, encefalitis y una enfermedad tipo roséola. Sin embargo, las presentaciones clínicas más comunes en niños con viremia HHV-7 son convulsiones y fiebre.

Los β -herpesvirus permanecen latentes en glándulas secretoras, células linforeticulares y en tejidos como los riñones, entre otros. Las muestras de plasma, suero, sangre total, células mononucleares de sangre periférica (PBMC), saliva y líquido cefalorraquídeo (LCR) son los tipos de muestras más aceptadas para la detección y el monitorización mediante PCR de HHV-6 y HHV-7.

El herpesvirus humano 8 (HHV-8) es un γ -herpesvirus asociado al sarcoma de Kaposi. Se transmite durante el contacto íntimo, probablemente a través de la transferencia de secreciones corporales como la saliva y las secreciones genitales. HHV-8 también se puede transmitir a través de transfusiones de sangre y trasplante de órganos. En niños inmunocompetentes, la infección primaria por HHV-8 se asocia con síntomas leves e inespecíficos de fiebre y erupción cutánea. En adultos, la infección primaria HHV-8 se ha relacionado con



enfermedades clínicas caracterizadas por diarrea, fatiga y linfadenopatía. Después de la resolución de la infección primaria, HHV-8 entra en un estado de latencia en las células de origen linfoide. Biopsias a partir de lesiones orales y/o cutáneas, saliva, sangre, plasma y PBMC son las muestras clínicas más frecuentemente utilizadas para la detección viral y la monitorización del tratamiento del HHV-8.

Las infecciones por HHV-6 y HHV-7 son comunes en los receptores de trasplantes (ej. trasplante de órgano sólido (TOS)), pero las enfermedades clínicas debidas a estos virus son poco frecuentes. Fiebre, erupción cutánea y supresión de la médula ósea son las manifestaciones más comunes, pero se pueden observar otros síntomas por una enfermedad invasiva de tejidos circundantes. Las infecciones debidas a HHV-8 pueden provocar sarcoma de Kaposi y, con menor frecuencia, enfermedad de Castleman, linfomas de efusión primaria, supresión de la médula ósea y fallo multiorgánico. La enfermedad HHV-8 es más común entre los receptores de trasplantes de regiones endémicas, aunque ciertas poblaciones en regiones no endémicas también tienen un alto riesgo.

La amplificación del DNA viral mediante la reacción en cadena de la polimerasa se usa ampliamente en todo el mundo por su sensibilidad y especificidad. qPCR es una técnica ideal para controlar las infecciones actuales.

3. Procedimiento

VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit está diseñado para el diagnóstico de HHV-6, HHV-7 y HHV-8 en muestras clínicas. Tras el aislamiento del DNA, la identificación de Herpesvirus humano 6, Herpesvirus humano 7 y Herpesvirus humano 8 se realiza mediante la reacción en cadena de la polimerasa utilizando oligonucleótidos específicos y una sonda marcada con fluorescencia que hibridan en una región conservada de los genes *U65-U66* para Herpesvirus humano 6, *ORF73* para Herpesvirus humano 8 y *U57* para Herpesvirus humano 7.

VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit aprovecha la actividad 5' exonucleasa de la DNA-polimerasa. Durante la amplificación del DNA, esta enzima hidroliza la sonda unida a la secuencia de DNA complementaria, separando el fluoróforo del *quencher*. Esta reacción genera un aumento en la señal fluorescente proporcional a la cantidad de DNA diana. Esta fluorescencia se puede monitorizar en equipos de PCR a tiempo real.

VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit contiene en cada pocillo todos los componentes necesarios para llevar a cabo la PCR a tiempo real (cebadores/sondas específicos, dNTPS, tampón, polimerasa) en formato estabilizado, así como un control interno para descartar la inhibición de la actividad polimerasa. Tras la reacción de amplificación Herpesvirus humano 6 se detecta en el canal FAM, Herpesvirus humano 8 se detecta en el canal ROX, Herpesvirus humano 7 se detecta en el canal Cy5 y el control interno(CI) se detecta en el canal HEX, VIC o JOE (Seleccionar el canal de detección apropiado según el equipo utilizado, ver Anexo 2).

4. Reactivos suministrados

VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit incluye los siguientes materiales y reactivos detallados en las Tablas 1, 2 y 3. Basado en la presentación comercial y la plataforma de PCR en tiempo real utilizada, la mezcla de reacción de PCR estabilizada se puede encontrar en diferentes tubos o pocillos y por



tanto comercializar en múltiples formatos. La Tabla 1 incluye materiales y reactivos para usar con dispositivos compatibles para tiras de 8 pocillos (Ver Anexo 1). La Tabla 2 incluye materiales y reactivos para usar con dispositivos compatibles para placas de 96 pocillos (Ver Anexo 1). La Tabla 3 incluye materiales y reactivos para usar con los instrumentos Qiagen / Corbett Rotor-Gene® para tiras de 4 pocillos.

Reactivo/Material	Descripción	Color	Cantidad
Human Herpes Virus 6, 7 & 8 8-well strips	Una mezcla de enzimas, cebadores-sondas, tampón, dNTPs, estabilizadores y Control interno en formato estabilizado	Blanco	6/12 tiras de 8 pocillos
Rehydration Buffer	Solución para la reconstitución del producto estabilizado	Azul	1 vial x 1.8 mL
Human Herpes Virus 6, 7 & 8 Positive Control	DNA sintético liofilizado no infeccioso	Rojo	1 vial
Negative control	Control negativo	Morado	1 vial x 1 mL
Water RNase/DNase free	Agua libre de RNase/DNase	Blanco	1 vial x 1 mL
Tear-off 8-cap strips	Tapones ópticos para sellar los pocillos durante el ciclo térmico	Transparente	6/12 tiras de 8 tapones

Tabla 1. Reactivos y materiales proporcionados en VIASURE Human Herpes Virus 6, 7 & 8 Real Time PCR Detection Kit con Ref. VS-HHV106L, VS-HHV106H, VS-HHV112L y VS-HHV112H.

Reactivo/Material	Descripción	Color	Cantidad
Human Herpes Virus 6, 7 & 8 96-well plate	Una mezcla de enzimas, cebadores-sondas, tampón, dNTPs, estabilizadores y Control interno en formato estabilizado	Blanco	1 placa
Rehydration Buffer	Solución para la reconstitución del producto estabilizado	Azul	1 vial x 1.8 mL
Human Herpes Virus 6, 7 & 8 Positive Control	DNA sintético liofilizado no infeccioso	Rojo	1 vial
Negative control	Control negativo	Morado	1 vial x 1 mL
Water RNase/DNase free	Agua libre de RNase/DNase	Blanco	1 vial x 1 mL
Tear-off 8-cap strips	Tapones ópticos para sellar los pocillos durante el ciclo térmico	Transparente	12 tiras de 8 tapones

Tabla 2. Reactivos y materiales proporcionados en VIASURE Human Herpes Virus 6, 7 & 8 Real Time PCR Detection Kit con Ref. VS-HHV113L y VS-HHV113H.

Reactivo/Material	Descripción	Color	Cantidad
Human Herpes Virus 6, 7 & 8 4-well strips	Una mezcla de enzimas, cebadores-sondas, tampón, dNTPs, estabilizadores y Control interno en formato estabilizado	Transparente	9/18 tiras de 4 pocillos
Rehydration Buffer	Solución para la reconstitución del producto estabilizado	Azul	1 vial x 1.8 mL
Human Herpes Virus 6, 7 & 8 Positive Control	DNA sintético liofilizado no infeccioso	Rojo	1 vial
Negative control	Control negativo	Morado	1 vial x 1 mL
Water RNase/DNase free	Agua libre de RNase/DNase	Blanco	1 vial x 1 mL
4-cap strips	Tapones ópticos para sellar los pocillos durante el ciclo térmico	Transparente	9/18 tiras de 4 tapones

Tabla 3. Reactivos y materiales proporcionados en VIASURE Human Herpes Virus 6, 7 & 8 Real Time PCR Detection Kit con Ref. VS-HHV136 y VS-HHV172. Para usar con instrumentos Qiagen / Corbett Rotor-Gene® y accesorios compatibles con tiras de 4 tubos 0.1 mL (72-Well Rotor y Locking Ring 72-Well Rotor).



5. Material requerido y no suministrado

La siguiente lista incluye los materiales que se requieren para el uso pero que no se incluyen en VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit.

- Equipo de PCR a tiempo real (termociclador).
- Kit de extracción de DNA.
- Centrífuga para tubos de 1.5 mL y para tiras de tubos de PCR o placas de 96 pocillos (si está disponible).
- Vórtex.
- Micropipetas (0.5-20 µL, 20-200 µL).
- Puntas con filtro.
- Guantes desechables sin polvo.
- Loading block (para usar con e instrumentos Qiagen/Corbett Rotor-Gene®).

VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit ha sido validado en los siguientes equipos: Applied Biosystems 7500 Fast Real-Time PCR System, Bio-Rad CFX96™ Real-Time PCR Detection System, Agilent Technologies AriaMx Real-Time PCR System, DNA-Technology DTprime Real-time Detection Thermal Cycler, DNA-Technology DTLite Real-time Detection Thermal Cycler, Rotor-Gene® Q (Qiagen), SmartCycler® (Cepheid), Roche Molecular Diagnostics Cobas z480 Analyzer, VIASURE 48 Real Time PCR System y VIASURE 96 Real Time PCR System. Cuando se utiliza el equipo Applied Biosystems 7500 Fast con tiras, se recomienda colocar el soporte adecuado para reducir el riesgo de aplastar el tubo (Ref. PN 4388506).

Para verificar la compatibilidad de los termocicladores, consulte el Anexo 1, para verificar los canales de detección más comunes, consulte el Anexo 2 y para verificar la configuración de la exposición de medición óptica, ver Anexo 3.

6. Condiciones de transporte y almacenamiento

- El transporte y almacenaje de los kits puede realizarse de 2-40°C hasta la fecha de caducidad indicada en la etiqueta.
- Almacenar el control positivo a -20°C tras su re-suspensión. Se recomienda separar en alícuotas para minimizar los ciclos de congelación y descongelación. Se ha validado la estabilidad del control positivo tras 6 ciclos de congelación y descongelación.
- Proteger los componentes de la luz.

7. Precauciones para el usuario

- El producto está destinado para uso exclusivo de usuarios profesionales, como profesionales o técnicos de laboratorio y sanitarios, entrenados en técnicas de biología molecular.
- No se recomienda usar el kit después de la fecha de caducidad.
- No utilizar los reactivos si los sobres o las bolsas que protegen los tubos están abiertos o dañados en el momento que se reciben.



- No utilizar los tubos de reacción si el material desecante que se incluye en cada sobre de aluminio no está o está dañado.
- No retirar el material desecante de los sobres de aluminio que contienen los tubos de reacción una vez abiertos.
- Cerrar los sobres de aluminio que protegen los tubos de reacción con el cierre zip inmediatamente después de cada uso (si está disponible, Ref. VS-HHV13L, VS-HHV113H, VS-HHV136 y VS-HHV172). Antes de cerrar los sobres eliminar cualquier exceso de aire.
- No utilizar los tubos de reactivos si el aluminio protector está roto o dañado.
- No mezclar reactivos de diferentes sobres y/o kits y/o lotes y/u otro proveedor.
- Proteger los reactivos de la humedad. Una exposición prolongada a la humedad puede afectar al rendimiento del producto.
- Para referencias VS-HHV136 y VS-HHV172 (compatible con instrumentos Qiagen/Corbett Rotor-Gene®) utilice el loading block para pipetear reactivos y muestras en cada tubo y para ayudar en el ajuste correcto de las tapas así como para evitar la contaminación.
- Diseñar un flujo de trabajo unidireccional. Se debe comenzar en el área de extracción y después pasar al área de amplificación y de detección. No poner en contacto las muestras, equipos y reactivos utilizados en un área con la zona en la que se realizó el paso anterior.
- Seguir las Buenas Prácticas de Laboratorio. Use ropa protectora, guantes de uso desechables, gafas y mascarilla. No comer, beber o fumar en el área de trabajo. Una vez terminada la prueba, lavarse las manos.
- Las muestras deben ser tratadas como potencialmente infecciosas, así como los reactivos que han estado en contacto con las muestras y deben ser gestionadas según la legislación sobre residuos sanitarios nacional. Tome las precauciones necesarias durante la recogida, almacenamiento, tratamiento y eliminación de muestras.
- Se recomienda la descontaminación periódica de los equipos usados habitualmente, especialmente micropipetas, y de las superficies de trabajo.
- Consulte las hojas de seguridad, previa solicitud.
- Consulte el manual de cada equipo de PCR a tiempo real para advertencias adicionales, precauciones y procedimientos.

8. Procedimiento del test

8.1. Extracción de DNA

Realizar la preparación de la muestra de acuerdo con las recomendaciones que aparecen en las instrucciones de uso del kit de extracción utilizado.

Para la extracción de DNA a partir de muestras clínicas puede utilizar su sistema optimizado de rutina manual o automático. Además, se puede usar cualquier kit de extracción de DNA disponible en el mercado y seguir las instrucciones de uso del fabricante. Los siguientes kits de extracción han sido validados:

- Viasure RNA-DNA Extraction kit (VIASURE), recomendado.
- Invisorb® Spin Universal Kit (Stratec).



8.2. Control positivo liofilizado

El vial de *Human Herpes Virus 6, 7 & 8 Positive Control* contiene una gran cantidad de copias molde por lo que se recomienda abrirlo y manipularlo en una zona del laboratorio separada del resto de los componentes. Reconstituir *Human Herpes Virus 6, 7 & 8 Positive Control* liofilizado (vial rojo) añadiendo 100 µL de Agua libre de RNAsa/DNAsa (vial blanco) suministrada y mezclar bien con la ayuda del vórtex. Almacenar el control positivo a -20°C tras su re-suspensión. Se recomienda separar en alícuotas para minimizar los ciclos de congelación y descongelación.

8.3. Protocolo PCR

Determinar y separar el número de reacciones necesarias incluyendo las muestras y los controles. En cada serie de muestras para cada uno de los ensayos a analizar se deben incluir un control positivo y uno negativo. Retirar el aluminio protector de las placas o tiras.

- 1) Reconstituir el número de pocillos que sean necesarios.

Añadir 15 µL del Rehydration Buffer (vial azul) en cada pocillo.

- 2) Añadir muestras y controles.

Añadir 5 µL de DNA extraído de cada muestra, de *Human Herpes Virus 6, 7 & 8 Positive Control* reconstituido (vial rojo) o Negative Control (vial morado) y cerrar los pocillos con los tapones suministrados.

Se recomienda centrifugar brevemente las tiras de 8 pocillos o las placas de 96 pocillos, o golpear suavemente cada tira sobre una superficie dura para asegurarse de que todos los líquidos queden en el fondo de los tubos (para los kits compatible con Qiagen/Corbett Rotor-Gene®).

Colocar la placa o las tiras en el termociclador.

- 3) Configurar el termociclador (para verificar la compatibilidad, consulte el Anexo 1).

Programar el termociclador siguiendo las condiciones descritas en la siguiente tabla e iniciar el programa:

Ciclos	Etapas	Tiempo	Temperatura
1	Desnaturalización inicial	2 min	95°C
45	Desnaturalización	10 seg	95°C
	Hibridación/Elongación (Recogida de datos*)	50 seg	60°C

Tabla 4. Protocolo PCR

Los datos de fluorescencia deben recogerse durante la etapa de elongación (*) a través de los canales FAM (Herpesvirus humano 6), ROX (Herpesvirus humano 8), Cy5 (Herpesvirus humano 7) y HEX, JOE o VIC (Control Interno). Dependiendo del equipo a utilizar seleccionar el canal de detección adecuado (ver Anexo 2). En los termocicladores Applied Biosystems 7500 Fast Real-Time PCR System y Stratagene Mx3005P™ Real Time PCR System comprobar que la opción del control pasivo ROX está desactivada. En el termociclador Applied



Biosystems 7500 Fast Real-Time PCR System seleccionar Ramp Speed Standard en el menú Select New Experiment/Advanced Setup/Experiment Properties.

9. Interpretación de resultados

El uso de los controles positivo y negativo junto con cada serie de muestras a analizar, valida la reacción comprobando la ausencia de señal en el pocillo del control negativo y la presencia de una señal en el pocillo de control positivo de herpesvirus humano 6, herpesvirus humano 7 y herpesvirus humano 8. Comprobar la emisión de la señal del control interno para verificar el correcto funcionamiento de la mezcla de amplificación. El análisis de las muestras se realiza con el software propio del equipo de PCR a tiempo real de acuerdo con las instrucciones de uso del fabricante. Con ayuda de la siguiente tabla, leer y analizar los resultados:

Herpesvirus humano 6 (FAM)	Herpesvirus humano 8 (ROX)	Herpesvirus humano 7 (Cy5)	Control Interno (HEX)	Control Negativo	Control Positivo	Interpretación
+	+	+	+/-	-	+	Herpesvirus humano 6, Herpesvirus humano 8 y Herpesvirus humano 7 Positivos
-	-	-	+	-	+	Herpesvirus humano 6, Herpesvirus humano 8 y Herpesvirus humano 7 Negativos
+	-	-	+/-	-	+	Herpesvirus humano 6 Positivo, Herpesvirus humano 8 y Herpesvirus humano 7 Negativos
+	+	-	+/-	-	+	Herpesvirus humano 6 y Herpesvirus humano 8 Positivos, y Herpesvirus humano 7 Negativo
+	-	+	+/-	-	+	Herpesvirus humano 6 y Herpesvirus humano 7 Positivos, y Herpesvirus humano 8 Negativo
-	+	-	+/-	-	+	Herpesvirus humano 8 Positivo, Herpesvirus humano 6 y Herpesvirus humano 7 Negativos
-	+	+	+/-	-	+	Herpesvirus humano 8 y Herpesvirus humano 7 Positivos, Herpesvirus humano 6 Negativo
-	-	+	+/-	-	+	Herpesvirus humano 7 Positivo, Herpesvirus humano 6 y Herpesvirus humano 8 Negativos
-	-	-	-	-	+	Inválido
+	+	+	+	+	-	Inválido

Tabla 5. Interpretación

+: curva de amplificación

-: sin curva de amplificación


Una muestra se considera positiva, si el valor Ct obtenido es menor de 40 y el control interno muestra o no una gráfica de amplificación. En ocasiones, la detección del control interno no es necesaria, ya que la presencia de



un alto número inicial de copias del ácido nucleico diana puede causar una amplificación preferencial de esta última.

Una muestra se considera negativa, si no se detecta una curva de amplificación por encima del valor umbral, y el control interno si la presenta. La inhibición de la reacción de PCR puede ser excluida por la amplificación del control interno.

Figura 1. Ejemplo de gráficas de amplificación del control negativo y positivo. Experimento realizado en el equipo Bio-Rad CFX96™ Real-Time PCR Detection System.



El resultado se considera inválido si se observa una gráfica de amplificación en el control negativo o ausencia de señal en el pocillo del control positivo. En ese caso, se recomienda repetir el ensayo.

En caso de ausencia de la señal de control interno en los pocillos de muestra, se recomienda repetir el ensayo diluyendo la muestra 1:10 o repetir la extracción para descartar posibles problemas de inhibición.

En el caso de obtener un resultado de dudosa interpretación, se recomienda verificar la correcta realización de cada uno de los pasos y revisar los parámetros y la forma sigmoidea de la curva. Si la situación no se resuelve, se recomienda repetir el ensayo, preferiblemente por duplicado. El resultado de la prueba debe ser evaluado en el contexto del historial médico, los síntomas clínicos y otras pruebas de diagnóstico por un profesional de la salud.

10. Limitaciones del test

- El resultado de la prueba debe ser evaluado en el contexto del historial médico, los síntomas clínicos y otras pruebas de diagnóstico por un profesional de la salud.
- Este ensayo se podría utilizar con diferentes tipos de muestras, aunque sólo ha sido validado con muestras de cultivo, medio de transporte, sangre y LCR.
- El correcto funcionamiento de la prueba depende de la calidad de la muestra; el ácido nucleico deber ser extraído de forma adecuada de las muestras clínicas. Una forma inadecuada de recolección, almacenaje y/o transporte de las muestras puede dar lugar a falsos negativos.
- Se puede detectar un bajo número de copias molde diana por debajo del límite de detección, pero los resultados pueden no ser reproducibles.



- Existe la posibilidad de falsos positivos debido a la contaminación cruzada con Herpesvirus humano 6, Herpesvirus humano 7 y Herpesvirus humano 8 ya sea por muestras que contienen altas concentraciones de DNA molde diana o por contaminación por arrastre a partir de productos de PCR de reacciones anteriores.

11. Control de calidad

VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit contiene controles positivo y negativo que deben ser incluidos en cada ensayo para interpretar correctamente los resultados. Además, el control interno (CI) en cada pocillo confirma el correcto funcionamiento de la técnica.

12. Características del test

12.1. Sensibilidad y especificidad clínica

VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit se evaluó mediante paneles EQA diferentes (QCMD). Estos paneles constan de 20 muestras de plasma congelado o muestras de medio de transporte. Los resultados se compararon con los informes finales de los programas de QCMD. 9 de 10 muestras del panel del Virus del Herpes Humano 6 fueron positivas para el patógeno y 1 de 10 muestras del Panel de Trasplante resultaron positivas al HHV-6 utilizando VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit.

Los resultados muestran una alta sensibilidad y especificidad para detectar Herpesvirus humano 6 utilizando VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit.

12.2. Sensibilidad analítica

VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit tiene un límite de detección de ≥ 10 copias de DNA por reacción para Herpesvirus humano 6, Herpesvirus humano 7 y Herpesvirus humano 8. (Figura 2, 3 y 4).

Figura 2. Diluciones seriadas de un estándar de Herpesvirus Humano 6 (10^7 - 10^1 copias/reacción). Experimento realizado en el equipo Bio-Rad CFX96™ Real-Time PCR Detection System (canal FAM).




Figura 3. Diluciones seriadas de un estándar de Herpesvirus Humano 7 (10^7 - 10^1 copias/reacción). Experimento realizado en el equipo Bio-Rad CFX96™ Real-Time PCR Detection System (canal Cy5).





Figura 4. Diluciones seriadas de un estándar de Herpesvirus Humano 8 (10^7 - 10^1 copias/reacción). Experimento realizado en el equipo Bio-Rad CFX96™ Real-Time PCR Detection System (canal ROX).



12.3. Especificidad analítica

La especificidad del ensayo de *Human Herpes Virus 6, 7 & 8* fue confirmada probando un panel compuesto por diferentes microorganismos que representan los patógenos que representan los patógenos más comunes que pueden causar meningitis y afectan a los pacientes inmunosupresores en particular. No se detectaron reacciones cruzadas con casi ninguno de los siguientes microorganismos testados, excepto con los patógenos diana que detecta cada ensayo.



Prueba de reacción cruzada					
<i>E. coli</i> 0.1285;O18:H7:K1	-	<i>Streptococcus pneumoniae</i> Z022	-	Virus St Louis Encephalitis cepa 17D	-
<i>Listeria innocua</i> Serotipo 6a/cepa CCUG 15531	-	<i>Plasmodium falciparum</i> 3D7	-	Virus Tick-Borne encephalitis cepa Neudorfl	-
<i>Listeria ivanovii</i> Serovar 5/cepa CCUG 15528	-	<i>Cryptococcus gatti</i> Z156	-	Virus Chikungunya (S27 Petersfield)	-
<i>Listeria monocytogenes</i> Serotipo 1/2b	-	<i>Haemophilus influenzae</i> MinnA	-	Dengue 1 (cepa Hawaii A)	-
<i>Listeria monocytogenes</i> Serovar 4b/cepa CIP 59.53	-	<i>Neisseria meningitidis</i> serogrupo A	-	Dengue 2 (cepa New Guinea C)	-
<i>Bacteroides fragilis</i>	-	<i>Treponema pallidum</i>	-	Dengue 3 (H87)	-
<i>Pseudomonas aeruginosa</i>	-	<i>Toxoplasma gondii</i> Tipo II	-	Dengue 4 (H241)	-
<i>Yersinia enterocolitica</i> O:3	-	Enterovirus tipo 68	-	Virus West Nile (Heja)	-
<i>Klebsiella oxytoca</i>	-	Enterovirus tipo 71	-	Virus West Nile (NY99)	-
<i>Enterobacter cloacae</i> Serotipo Cloaca A	-	Virus Epstein-Barr	-	Virus West Nile (Ug37)	-
<i>Enterobacter aerogenes</i> Serotipo Cloaca B	-	Virus BK Tipo Ib2	-	Virus fiebre amarilla (cepa 17D)	-
<i>Enterococcus durans</i>	-	Virus BK Tipo IV	-	Echovirus Tipo 30	-
<i>Enterococcus faecalis</i>	-	Virus JC Tipo 2B	-	Echovirus Tipo 11	-
<i>Enterococcus faecium</i> serotipo 11	-	Virus JC Tipo 1A	-	HSV-1 cepa MacIntyre	-
<i>Acinetobacter baumannii</i>	-	Citomegalovirus cepa AD-169	-	HSV-2 MS	-
<i>Candida albicans</i>	-	Coxsackievirus tipo A24	-	HHV6 cepa Z29	-/+
<i>Aspergillus fumigatus</i>	-	Coxsackievirus tipo A9	-	HHV6 Tipo A	-/+
<i>Pneumocystis jirovecii</i> Tipo A1	-	Coxsackievirus tipo B3	-	HHV6 Tipo B	-/+
<i>Pneumocystis jirovecii</i> g885652	-	Parechovirus Tipo 3	-	Virus Varicella-Zoster Ellen	-
<i>Staphylococcus aureus</i> subsp <i>aureus</i>	-	Parvovirus B19	-	Virus Varicella Zoster OKA	-
<i>Streptococcus agalactiae</i> Z019	-				

Tabla 6. Microorganismos patógenos de referencia utilizados en este estudio.

12.4. Reactividad analítica

La reactividad de VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit para el Herpesvirus humano 6 se evaluó frente a HHV6 cepa Z29, HHV6 Tipo A y HHV6 Tipo B, mostrando un resultado positivo.

La reactividad de VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit para el Herpesvirus humano 8 se evaluó frente a una secuencia sintética de Herpesvirus humano 8, mostrando un resultado positivo.

La reactividad de VIASURE *Human Herpes Virus 6, 7 & 8* Real Time PCR Detection Kit para el Herpesvirus humano 7 se evaluó frente a una secuencia sintética de Herpesvirus humano 7, mostrando un resultado positivo.

13. Bibliography/Bibliografía

1. C.C.Y. Yip *et al.* Comparative evaluation of a laboratory developed real-time PCR assay and the RealStar® HHV-6 PCR Kit for quantitative detection of human herpesvirus 6. *Journal of Virological Methods* 2017; 112–116.



2. D.J. Speicher *et al.* Detection of human herpesvirus 8 by quantitative polymerase chain reaction: development and standardisation of methods. *BMC Infectious Diseases* 2012; 12:210.
3. K. Wada *et al.* Multiplex real-time PCR for the simultaneous detection of herpes simplex virus, human herpesvirus 6, and human herpesvirus 7. *Microbiology and Immunology* 2009; 22-29.
4. J.Carter *et al.* Virology, Principles and Applications. *John Wiley & Sons Ltd* 2007.
5. A. Gautheret-Dejean *et al.* Development of a real-time polymerase chain reaction assay for the diagnosis of human herpesvirus-6 infection and application to bone marrow transplant patients. *Journal of Virological Methods* 2002; 27-35.
6. R.R. Razonable *et al.* Human herpesviruses 6, 7 and 8 in solid organ transplant recipients. *American Journal of Transplantation* 2013; 67-77.

14. Symbols for IVD components and reagents/Símbolos para reactivos y productos para diagnóstico in vitro

	<i>In vitro</i> diagnostic device Producto para diagnóstico <i>in vitro</i>		Keep dry Almacenar en lugar seco		Use by Fecha de caducidad		Manufacturer Fabricante		Batch code Número de lote
	Consult instructions for use Consultar las instrucciones de uso		Temperature limitation Limitación de temperatura		Contains sufficient for <n> test Contiene <n> test		Sample diluent Diluyente de muestra		Catalogue number Número de referencia



ANEXO 1

COMPATIBILIDAD DE LOS EQUIPOS A TIEMPO REAL MÁS COMUNES

Las tiras de bajo perfil pueden usarse en todos los termocicladores equipados con un bloque de perfil bajo, como los sistemas listados en la tabla A.1. Las tiras de perfil alto pueden usarse en todos los termocicladores PCR equipados con bloque de perfil alto o normal (high profile), como los sistemas listados en la tabla A.2. Si no encuentra su termociclador en la siguiente lista, por favor póngase en contacto con su proveedor.

Tabla A.1 TERMOCICLADORES CON BLOQUE DE BAJO PERFIL	
Fabricante	Modelo
Agilent Technologies	AriaMx/AriaDx Real-Time PCR System
Applied Biosystems	7500 Fast Real-Time PCR System ⁽¹⁾
Applied Biosystems	7500 Fast Dx Real-Time PCR System ⁽¹⁾
Applied Biosystems	QuantStudio™ 12K Flex 96-well Fast
Applied Biosystems	QuantStudio™ 6 Flex 96-well Fast
Applied Biosystems	QuantStudio™ 7 Flex 96-well Fast
Applied Biosystems	QuantStudio™ 3 Fast Real-Time PCR System ⁽⁵⁾
Applied Biosystems	QuantStudio™ 5 Fast Real-Time PCR System
Applied Biosystems	StepOne Plus™ Real-Time PCR System ⁽⁵⁾
Applied Biosystems	StepOne™ Real-Time PCR System ⁽⁵⁾
Applied Biosystems	ViiA™ 7 Fast Real-Time PCR System
BIONEER	Exicycler™ 96
Bio-Rad	CFX96™ / CFX96™ IVD Real-Time PCR Detection System
Bio-Rad	Mini Opticon™ Real-Time PCR Detection System ⁽⁶⁾
Cepheid	SmartCycler® ⁽³⁾
Qiagen	Rotor-Gene® Q ⁽³⁾
Roche	LightCycler® 480 Real-Time PCR System ⁽⁴⁾
Roche	LightCycler® 96 Real-Time PCR System ⁽⁴⁾
Roche	Cobas z480 Analyzer ⁽⁴⁾

Tabla A.2 TERMOCICLADORES CON BLOQUE DE PERFIL ALTO	
Fabricante	Modelo
Abbott	Abbott m2000 RealTime System
Applied Biosystems	7300 Real-Time PCR System ⁽⁵⁾
Applied Biosystems	7500 Real-Time PCR System
Applied Biosystems	7900 HT Real-Time PCR System ⁽⁵⁾
Applied Biosystems	ABI PRISM 7000 ⁽⁶⁾
Applied Biosystems	ABI PRISM 7700 ⁽⁵⁾
Applied Biosystems	QuantStudio™ 12K Flex 96-well
Applied Biosystems	QuantStudio™ 6 Flex 96-well
Applied Biosystems	QuantStudio™ 7 Flex 96-well
Applied Biosystems	QuantStudio™ 3 Real-Time PCR System ⁽⁵⁾
Applied Biosystems	QuantStudio™ 5 Real-Time PCR System
Applied Biosystems	ViiA™ 7 Real-Time PCR System
Analytik Jena Biometra	TOptical
Analytik Jena Biometra	qTOWER 2.0
BIONEER	Exicycler™ 96
Bio-Rad	CFX96™ / CFX96™ IVD Real-Time PCR Detection System
Bio-Rad	iCycler iQ™ Real-Time PCR Detection System
Bio-Rad	iCycler iQ™5 Real-Time PCR Detection System
Bio-Rad	MyiQ™ Real-Time PCR Detection System ⁽⁶⁾
Bio-Rad	MyiQ™2 Real-Time PCR Detection System ⁽⁶⁾
Cepheid	SmartCycler® ⁽³⁾
DNA-Technology	DTprime Real-time Detection Thermal Cycler ⁽²⁾
DNA-Technology	DTlite Real-Time PCR System ⁽²⁾
Eppendorf	Mastercycler™ep realplex
Qiagen	Rotor-Gene® Q ⁽³⁾
Stratagene / Agilent Technologies	Mx3000P™ Real Time PCR System
Stratagene / Agilent Technologies	Mx3005P™ Real Time PCR System
VIASURE	VIASURE 48 Real Time PCR System ⁽²⁾
VIASURE	VIASURE 96 Real Time PCR System ⁽²⁾

(1) Seleccionar Ramp Speed "Standard".

(2) Ver Anexo 3 para la configuración de los valores de exposición.

(3) El producto se debe reconstituir siguiendo el procedimiento adecuado (ver Procedimiento del test) y transvasar a los tubos específicos Rotor-Gene® Q o SmartCycler®.

(4) Se necesita un soporte especial que ajuste con estos equipos Roche de PCR a tiempo real.

(5) No lectura en canal Cy5.

(6) Lectura solo en canales FAM y HEX.

Tabla A1/A2. Equipos compatibles de PCR a tiempo real más comunes.



ANEXO 2

CANALES DE DETECCIÓN DE LOS EQUIPOS A TIEMPO REAL MÁS COMUNES

Los canales de fluorescencia de algunos de los termocicladores a tiempo real más comunes se especifican en la Tabla A3.

TERMOCICLADORES A TIEMPO REAL	CANAL VIASURE	CANAL DE DETECCIÓN	OBSERVACIONES
Bio-Rad CFX96™	FAM	FAM	Algunos pocillos pueden tener una deriva anormal de la fluorescencia durante los ciclos iniciales de la carrera, dando lugar a una línea ascendente no sigmoidea. Si ve este efecto, en el menú Setting, seleccione la opción Apply Fluorescence Drift Correction dentro de Baseline Settings para corregirlo.
	HEX	HEX	
	ROX	ROX	
	Cy5	Cy5	
ABI 7500 Applied Biosystems	FAM	FAM	Opción del control pasivo ROX desactivada. Algunos pocillos pueden tener una deriva anormal de la fluorescencia durante los ciclos iniciales de la carrera, dando lugar a una línea ascendente no sigmoidea. Si ve este efecto, por favor modifique la línea base (Baseline): Seleccione los valores para Start Cycle y End Cycle de forma que la línea base termine antes de comienzo la detección de un aumento significativo de la fluorescencia.
	HEX	VIC	
	ROX	ROX	
	Cy5	Cy5	
Roche Lightcycler®480II	FAM	465/510	Se requiere compensación de color
	HEX	533/580	
	ROX	533/610	
	Cy5	618/660	
Smartcycler® Cepheid	FAM	Channel 1	
	HEX	Channel 2	
	ROX	Channel 3	
	Cy5	Channel 4	
Abbott m2000rt	FAM	FAM	
	HEX	VIC	
	ROX	ROX	
	Cy5	Cy5	
Mx3000P™ Mx 3005P™ Stratagene	FAM	FAM	Opción del control pasivo ROX desactivada
	HEX	VIC	
	ROX	ROX	
	Cy5	Cy5	
AriaMx Agilent	FAM	FAM	
	HEX	HEX	
	ROX	ROX	
	Cy5	Cy5	
Rotor-Gene®Q Qiagen	FAM	Green	Durante la configuración de los canales (Channel Setup), presione el botón "Gain Optimisation" y después vaya a "Optimise Acquiring". La fluorescencia del apartado Target Sample Range tiene que estar entre 5 y 10 FI para cada canal. Además, marque la opción "Perform Optimisation Before 1st Acquisition".
	HEX	Yellow	
	ROX	Orange	
	Cy5	Red	
Mic Real Time PCR Cycler bms	FAM	Green	En el menú "Run Profile", introduzca los parámetros correctos para "Temperature Control" (Standard TAQ (v3)), Volume (20 ul) y el protocolo térmico apropiado. En la ventana "Cycling", seleccione la opción "Acquire on" para todos los canales haciendo click sobre ellos. Utilice los valores de "Gain" que aparecen por defecto para cada canal (Green = 3, Yellow = 10, Orange = 10, Red = 10).
	HEX	Yellow	
	ROX	Orange	
	Cy5	Red	

Tabla A3: Canales de detección de fluorescencia de diferentes equipos de PCR a Tiempo Real



ANEXO 3

CONFIGURACIÓN DE LOS VALORES DE EXPOSICIÓN

Los parámetros de exposición de algunos termocicladores deben ajustarse para su adecuación y correcto funcionamiento con los test "VIASURE Real Time PCR Detection Kits". Este ensayo ha sido validado con los siguientes valores de exposición:

- DTprime Real-time Detection Thermal Cycler (DNA-Technology) y VIASURE 96 Real Time PCR System (CerTest Biotec S.L.): canal FAM -500*, canal HEX - 1000, canal ROX - 1000 y canal Cy5 -1000.
- DTlite Real-Time PCR System (DNA-Technology) y VIASURE 48 Real Time PCR System (CerTest Biotec S.L.): canal FAM -500, canal HEX - 500, canal ROX - 500 y canal Cy5 – 500.

*Si el resultado en el canal FAM no es el esperado, no hay amplificaciones o se observa elevado ruido de fondo, por favor, baje los valores de exposición indicados anteriormente hasta 150.



- CFX™ and IQ5™ are registered trademarks of Bio-Rad Laboratories.
- ABI®, QuantStudio™ and ViiA™ are registered trademarks of Thermo Fisher Scientific Inc.
- LightCycler® is a registered trademark of Roche.
- Mx3000P™, Mx3005™ and AriaMx are registered trademarks of Agilent Technologies.
- Mastercycler™ is a registered trademark of Eppendorf.
- Rotor-Gene®Q is a registered trademark of Qiagen.
- SmartCycler® is a registered trademark of Cepheid.

Revision: May 2019





CerTest Biotec, S.L.

Pol. Industrial Río Gállego II · Calle J, Nº1
50840, San Mateo de Gállego, Zaragoza (Spain)
www.certest.es



VIASURE online

F-362 rev01

VIASURE



CE

IVD



Real Time PCR Detection Kits

CerTest
BIOTEC