

GeneSys Diluent Reagent

REF

01-03-0019

Fill 3.8L

Pour d'autres langues
Für andere Sprachen
Para otras lenguas
Per le altre lingue
Dla innych języków

Para outras línguas
Για τις άλλεςλώσεις
För andra språk
For andre språk



www.trinitybiotech.com

INTENDED USE

GeneSys Diluent Reagent is designed for use with the Trinity Biotech ultra2 GeneSys systems for the separation and quantitation of normal and variant hemoglobin species in blood. For *in vitro* diagnostic use only. **IVD**

SUMMARY AND EXPLANATION OF TEST

Hemoglobin Variants – Assessment of hemoglobin variants has proven useful in the identification of altered hemoglobin commonly associated with disorders called hemoglobinopathies.

Reagents are performance validated to assure accuracy and precision with the Trinity Biotech assays and systems for the screening for hemoglobin variants.

Reagent is ready for use.

STORAGE AND STABILITY



Store at ambient temperatures (2 – 28°C). Do not allow to freeze.

The reagent is stable until the expiration date indicated on the label when kept tightly closed and protected from extreme environmental conditions.



See the container label for the expiration date. **DO NOT USE** after the expiration date.

PRECAUTIONS

For *in vitro* diagnostic use only. Avoid skin contact. Consult the product SDS for safety information. This reagent is used in conjunction with blood testing equipment and warrants handling under universal precaution procedures for safety.

ORDERING INFORMATION

Reference No.	Item	Quantity
01-03-0019	GeneSys Diluent	3.8L



Manufacturer



European Conformity



Authorized Representative



Reference Number



Lot Number



For *in vitro* Diagnostic Use



Expiration



Temperature limitation




Consult Instructions for Use



Sodium Azide



Primus Corporation
D.B.A. Trinity Biotech
Kansas City, MO 64132
Tel. 1 800-325-3424
Fax: 1 816-361-1974



Trinity Biotech plc
Bray Co. Wicklow, Ireland
Tel. 353 1 2769800
Fax 353 1 2769888
www.trinitybiotech.com

Rev B
05/16