



# AIR & EMISSIONS

Matrices consisting of organic, inorganic, and particulate matter for testing emissions and ambient air. Standards are designed to meet regulations of the United States Environmental Protection Clean Air Act and may be used to satisfy PT requirements worldwide.

## Air & Emissions PT Schedule 2022

### Air & Emissions

|   | Scheme # | Opens  | Closes |
|---|----------|--------|--------|
| Q | AE 59    | Jan 31 | Mar 17 |
| Q | AE 60    | Apr 25 | Jun 9  |
| Q | AE 61    | Jul 29 | Sep 12 |
| Q | AE 62    | Oct 28 | Dec 12 |

## 2023

### Air & Emissions

|   | Scheme # | Opens  | Closes |
|---|----------|--------|--------|
| Q | AE 63    | Jan 30 | Mar 16 |
| Q | AE 64    | Apr 28 | Jun 12 |
| Q | AE 65    | Jul 28 | Sep 11 |
| Q | AE 66    | Oct 27 | Dec 11 |

## Contents

**CRM:** A reference material characterized by a metrologically valid procedure for one or more specified properties, accompanied by a reference material certificate that provides the value of the specified property, its associated uncertainty, and a statement of metrological traceability.

A complete listing of ERA's CRMs can be found on our Scope of Accreditation for general requirements for competence of reference material producers available at [www.eraqc.com/AboutERA/Accreditations](http://www.eraqc.com/AboutERA/Accreditations).

**PT:** A Proficiency Test (PT) is an analysis of what is often referred to as a blind sample or a sample with unknown concentrations of analytes for the purpose of evaluating a laboratory's analytical performance.

**QR:** Similar to a Proficiency Test, a QuiK Response (QR) is a sample with unknown concentrations. However, unlike a scheduled PT, QR is on-demand and available at any time. Plus, your results are returned within two business days. QuiK Response can be used as a bilateral PT as referenced in the IUPAC/CITAC guide: Selection and use of PT schemes for a limited number of participants – chemical analytical labs.

**RM:** A material, sufficiently homogeneous and stable with respect to one or more specified properties, which has been established to be fit for its intended use in a measurement process.

| Description   | CRM/<br>RM | PT            | QR     | Page |
|---|------------|---------------|--------|------|
| Aldehydes and Ketones on Sorbent                    | 1114       | 1014 <b>Q</b> | 1114QR | 55   |
| Ammonia in Impinger Solution                        | 1145       | 1045 <b>Q</b> | 1145QR | 57   |
| Chromium on Filter Paper                            | 1131       | 1031 <b>Q</b> | 1131QR | 56   |
| Fluoride in Impinger Solution                       | 1141       | 1041 <b>Q</b> | 1141QR | 57   |
| Hexavalent Chromium in Impinger Solution            | 1132       | 1032 <b>Q</b> | 1132QR | 56   |
| Hydrogen Halides & Halogens in Impinger Solution    | 1140       | 1040 <b>Q</b> | 1140QR | 57   |
| Lead in Impinger Solution                           | 1130       | 1030 <b>Q</b> | 1130QR | 56   |
| Lead on Filter Paper                                | 1129       | 1029 <b>Q</b> | 1129QR | 56   |
| Mercury in Impinger Solution                        | 1128       | 1028 <b>Q</b> | 1128QR | 56   |
| Mercury on Filter Paper                             | 1127       | 1027 <b>Q</b> | 1127QR | 56   |
| Metals on Filter Paper                              | 1125       | 1025 <b>Q</b> | 1125QR | 56   |
| Metals in Impinger Solution                         | 1126       | 1026 <b>Q</b> | 1126QR | 56   |
| Nitrogen Oxide in Impinger Solution                 | 1142       | 1042 <b>Q</b> | 1142QR | 57   |
| Organochlorine Pesticides on Polyurethane Foam      | 1111       | 1011 <b>Q</b> | 1111QR | 55   |
| PAHs on Polyurethane Foam                           | 1113       | 1013 <b>Q</b> | 1113QR | 55   |
| Particulate Matter in Impinger Solution             | 1151       | 1051 <b>Q</b> | 1151QR | 57   |
| Particulate Matter on Filter Paper                  | 1150       | 1050 <b>Q</b> | 1150QR | 57   |
| PCBs on Polyurethane Foam                           | 1112       | 1012 <b>Q</b> | 1112QR | 55   |
| Semivolatiles on Polyurethane Foam                  | 1110       | 1010 <b>Q</b> | 1110QR | 55   |
| Sulfur Dioxide in Impinger Solution                 | 1143       | 1043 <b>Q</b> | 1143QR | 57   |
| Sulfuric Acid & Sulfur Dioxide in Impinger Solution | 1144       | 1044 <b>Q</b> | 1144QR | 57   |
| Volatiles in Gas Cylinder                           | 1100       | 1000 <b>Q</b> | 1100QR | 54   |
| Volatiles on Sorbent                                | 1101       | 1001 <b>Q</b> | 1101QR | 54   |

**Q** All Waters ERA Air & Emissions PTs open quarterly. Quarterly months are January, April, July, and October.

# Volatiles

## Volatiles in Gas Cylinder\*

| RM**<br>Cat. #1100 | PT<br>Cat. #1000 | Q | QR<br>Cat. #1100QR |
|--------------------|------------------|---|--------------------|
|--------------------|------------------|---|--------------------|

One pressurized gas cylinder containing 87 L of gas at 1500 psig (103 bar) for use with EPA methods TO-14, TO-15, or other applicable methods. Contains at least 10 analytes, randomly selected from the list below, at 2-50 ppbv (4-100 ppbv) for Total Xylenes.

|                                    |                                |                           |
|------------------------------------|--------------------------------|---------------------------|
| Acetone                            | 1,1-Dichloroethane             | Styrene                   |
| Benzene                            | 1,2-Dichloroethane             | 1,1,2,2-Tetrachloroethane |
| Benzyl chloride                    | 1,1-Dichloroethylene           | Tetrachloroethylene       |
| Bromodichloromethane               | cis-1,2-Dichloroethylene       | Toluene                   |
| Bromoform                          | trans-1,2-Dichloroethylene     | Trichloroethene           |
| Bromomethane                       | 1,2-Dichloropropane            | 1,2,4-Trichlorobenzene    |
| 1,3-Butadiene                      | cis-1,3-Dichloropropylene      | 1,1,1-Trichloroethane     |
| 2-Butanone (MEK)                   | trans-1,3-Dichloropropylene    | 1,1,2-Trichloroethane     |
| Methyl tert-butyl ether (MTBE)     | 1,2-Dichlorotetrafluoroethane  | Trichlorofluoromethane    |
| Carbon disulfide                   | (Freon 114)                    | (Freon 11)                |
| Carbon tetrachloride               | Ethyl acetate                  | Trichlorotrifluoromethane |
| Chlorobenzene                      | Ethylbenzene                   | (Freon 113)               |
| Chlorodibromomethane               | p-Ethyltoluene                 | 1,2,4-Trimethylbenzene    |
| Chloroethane                       | n-Heptane                      | 1,3,5-Trimethylbenzene    |
| Chloroform                         | Hexachlorobutadiene            | Vinyl bromide             |
| Chloromethane                      | n-Hexane                       | Vinyl chloride            |
| Cyclohexane                        | 2-Hexanone                     | Xylenes, total            |
| 1,2-Dibromoethane (EDB)            | Isopropyl alcohol              | m&p-Xylene                |
| 1,2-Dichlorobenzene                | Methylene chloride             | o-Xylene                  |
| 1,3-Dichlorobenzene                | Methyl methacrylate            |                           |
| 1,4-Dichlorobenzene                | 4-Methyl-2-pentanone (MIBK)    |                           |
| Dichlorodifluoromethane (Freon 12) | Methyl tert-butyl ether (MTBE) |                           |
|                                    | Propylene                      |                           |

\*Volatiles in Gas Cylinder ships as dangerous goods.

\*\* Reference Material (RM)

## Volatiles on Sorbent

| CRM<br>Cat. #1101 | PT<br>Cat. #1001 | Q | QR<br>Cat. #1101QR |
|-------------------|------------------|---|--------------------|
|-------------------|------------------|---|--------------------|

One 2 mL flame-sealed ampule for spiking client-specific sorbent. Use with EPA Methods TO-17, 0030, 0031, or other applicable methods. Contains at least 24 analytes, randomly selected from the list below, at 50–2000 ng/sample (200–3000 ng/sample for Total Xylenes) after preparation.

|                           |                                    |                             |
|---------------------------|------------------------------------|-----------------------------|
| Acetone                   | 1,1-Dichloropropene                | Methylene chloride          |
| Acetonitrile              | 1,2-Dibromo-3-chloropropane (DBCP) | 4-Methyl-2-pentanone (MIBK) |
| Acrolein                  | 1,2-Dibromoethane (EDB)            | Naphthalene                 |
| Acrylonitrile             | Dibromomethane                     | Nitrobenzene                |
| Benzene                   | 1,2-Dichlorobenzene                | n-Propylbenzene             |
| Bromobenzene              | 1,3-Dichlorobenzene                | Styrene                     |
| Bromochloromethane        | 1,4-Dichlorobenzene                | 1,1,2-Tetrachloroethane     |
| Bromodichloromethane      | Dichlorodifluoromethane (Freon 12) | 1,1,2,2-Tetrachloroethane   |
| Bromoform                 | 1,1-Dichloroethane                 | Tetrachloroethene           |
| Bromomethane              | 1,2-Dichloroethane                 | Toluene                     |
| 2-Butanone (MEK)          | 1,2-Dichloroethane                 | 1,2,3-Trichlorobenzene      |
| n-Butylbenzene            | 1,1-Dichloroethene                 | 1,2,4-Trichlorobenzene      |
| sec-Butylbenzene          | cis-1,2-Dichloroethene             | 1,1,1-Trichloroethane       |
| tert-Butylbenzene         | trans-1,2-Dichloroethene           | 1,1,2-Trichloroethane       |
| Carbon disulfide          | 1,2-Dichloropropane                | Trichloroethylene           |
| Carbon tetrachloride      | cis-1,3-Dichloropropene            | Trichlorofluoromethane      |
| Chlorobenzene             | trans-1,3-Dichloropropene          | 1,2,3-Trichloropropane      |
| Chlorodibromomethane      | Ethylbenzene                       | 1,2,4-Trimethylbenzene      |
| Chloroethane              | Hexachlorobutadiene                | 1,3,5-Trimethylbenzene      |
| 2-Chloroethyl vinyl ether | Hexachloroethane                   | Vinyl acetate               |
| Chloroform                | 2-Hexanone                         | Vinyl chloride              |
| Chloromethane             | Isopropylbenzene                   | Xylenes, total              |
| 2-Chlorotoluene           | 4-Isopropyltoluene                 | m&p-Xylene                  |
| 4-Chlorotoluene           | Methyl tert-butyl ether (MTBE)     | o-Xylene                    |
| 1,3-Dichloropropane       |                                    |                             |
| 2,2-Dichloropropane       |                                    |                             |

## Semivolatiles

## Semivolatiles on Polyurethane Foam

| CRM<br>Cat. #1110 | PT<br>Cat. #1010 | Q | QR<br>Cat. #1110QR |
|-------------------|------------------|---|--------------------|
|-------------------|------------------|---|--------------------|

Two 2 mL flame-sealed ampules plus one polyurethane foam. Use with EPA Method 0010, or other applicable methods. Contains at least 42 analytes, randomly selected from the list below, at 10–225 µg/sample (200–1000 µg/sample for Benzidine) after preparation.

|                             |                           |                              |
|-----------------------------|---------------------------|------------------------------|
| Acenaphthene                | 1,3-Dichlorobenzene       | N-Nitroso-di-n-propylamine   |
| Acenaphthylene              | 1,4-Dichlorobenzene       | 2,2'-Oxybis(1-chloropropane) |
| Aniline                     | 3,3'-Dichlorobenzidine    | Pentachlorobenzene           |
| Anthracene                  | Diethyl phthalate         | Phenanthrene                 |
| Benzidine                   | Dimethyl phthalate        | Pyrene                       |
| Benzo(a)anthracene          | 2,4-Dinitrotoluene        | Pyridine                     |
| Benzo(b)fluoranthene        | 2,6-Dinitrotoluene        | o-Toluidine                  |
| Benzo(k)fluoranthene        | Di-n-octyl phthalate      | 1,2,4,5-Tetrachlorobenzene   |
| Benzo(g,h,i)perylene        | Fluoranthene              | 1,2,4-Trichlorobenzene       |
| Benzo(a)pyrene              | Fluorene                  | Benzoic Acid                 |
| Benzyl alcohol              | Hexachlorobenzene         | 4-Chloro-3-methylphenol      |
| 4-Bromophenyl phenyl ether  | Hexachlorobutadiene       | 2-Chlorophenol               |
| Butyl benzyl phthalate      | Hexachlorocyclopentadiene | 2,4-Dichlorophenol           |
| Carbazole                   | Hexachloroethane          | 2,6-Dichlorophenol           |
| 4-Chloroaniline             | Indeno(1,2,3-cd)pyrene    | 2,4-Dimethylphenol           |
| Bis(2-chloroethoxy)methane  | Isophorone                | 2,4-Dinitrophenol            |
| Bis(2-chloroethyl)ether     | Naphthalene               | 2-Methyl-4,6-dinitrophenol   |
| Bis(2-ethylhexyl)phthalate  | 2-Methylnaphthalene       | 2-Methylphenol (o-Cresol)    |
| 1-Chloronaphthalene         | Naphthalene               | 4-Methylphenol (p-Cresol)    |
| 2-Chloronaphthalene         | 2-Nitroaniline            | 2-Nitrophenol                |
| 4-Chlorophenyl phenyl ether | 3-Nitroaniline            | 4-Nitrophenol                |
| Chrysene                    | 4-Nitroaniline            | Pentachlorophenol            |
| Dibenz(a,h)anthracene       | Nitrobenzene              | Phenol                       |
| Dibenzofuran                | N-Nitrosodiethylamine     | 2,3,4,6-Tetrachlorophenol    |
| Di-n-butyl phthalate        | N-Nitrosodimethylamine    | 2,4,5-Trichlorophenol        |
| 1,2-Dichlorobenzene         | (NDMA)                    | 2,4,6-Trichlorophenol        |
|                             | N-Nitrosodiphenylamine    |                              |

## Organochlorine Pesticides on Polyurethane Foam

| CRM<br>Cat. #1111 | PT<br>Cat. #1011 | Q | QR<br>Cat. #1111QR |
|-------------------|------------------|---|--------------------|
|-------------------|------------------|---|--------------------|

One 2 mL flame-sealed ampule plus one polyurethane foam. Use with EPA Methods TO-04A, TO-10A, or other applicable methods. Contains at least 16 analytes, randomly selected from the list below, at 1–20 µg/sample after preparation.

|                     |                    |                           |
|---------------------|--------------------|---------------------------|
| Aldrin              | 4,4'-DDD           | Endrin                    |
| alpha-BHC           | 4,4'-DDE           | Endrin aldehyde           |
| beta-BHC            | 4,4'-DDT           | Endrin ketone             |
| delta-BHC           | Dieldrin           | Heptachlor                |
| gamma-BHC (Lindane) | Endosulfan I       | Heptachlor epoxide (beta) |
| alpha-Chlordane     | Endosulfan II      | Methoxychlor              |
| gamma-Chlordane     | Endosulfan sulfate |                           |

## PCBs on Polyurethane Foam

| CRM<br>Cat. #1112 | PT<br>Cat. #1012 | Q | QR<br>Cat. #1112QR |
|-------------------|------------------|---|--------------------|
|-------------------|------------------|---|--------------------|

One 2 mL flame-sealed ampule plus one polyurethane foam. Use with EPA Methods TO-04A, TO-10A, or other applicable methods. Contains one aroclor, randomly selected from the list below, at 2–10 µg/sample after preparation.

|              |              |              |
|--------------|--------------|--------------|
| Aroclor 1016 | Aroclor 1242 | Aroclor 1260 |
| Aroclor 1221 | Aroclor 1248 |              |
| Aroclor 1232 | Aroclor 1254 |              |

## PAHs on Polyurethane Foam

| CRM<br>Cat. #1113 | PT<br>Cat. #1013 | Q | QR<br>Cat. #1113QR |
|-------------------|------------------|---|--------------------|
|-------------------|------------------|---|--------------------|

One 2 mL flame-sealed ampule plus one polyurethane foam. Use with EPA Method TO-13A, or other applicable methods. Contains at least 13 analytes, randomly selected from the list below, at 10–200 µg/sample after preparation.

|                      |                       |                        |
|----------------------|-----------------------|------------------------|
| Acenaphthene         | Benzo(g,h,i)perylene  | Indeno(1,2,3-cd)pyrene |
| Acenaphthylene       | Benzo(a)pyrene        | 1-Methylnaphthalene    |
| Anthracene           | Chrysene              | 2-Methylnaphthalene    |
| Benzo(a)anthracene   | Dibenz(a,h)anthracene | Naphthalene            |
| Benzo(b)fluoranthene | Fluoranthene          | Phenanthrene           |
| Benzo(k)fluoranthene | Fluorene              | Pyrene                 |

## Aldehydes &amp; Ketones on Sorbent

| CRM<br>Cat. #1114 | PT<br>Cat. #1014 | Q | QR<br>Cat. #1114QR |
|-------------------|------------------|---|--------------------|
|-------------------|------------------|---|--------------------|

One 2 mL flame-sealed ampule to be spiked onto sorbent. Use with EPA Method TO-11A, or other applicable methods. Contains at least four analytes, randomly selected from the list below, at 0.5–10 µg/sample after preparation.

|                         |                          |                            |
|-------------------------|--------------------------|----------------------------|
| Acetaldehyde            | Crotonaldehyde           | Propionaldehyde (Propanal) |
| Acetone                 | 2,5-Dimethylbenzaldehyde | o-Tolualdehyde             |
| Benzaldehyde            | Formaldehyde             | m-Tolualdehyde             |
| 2-Butanone (MEK)        | Hexaldehyde (Hexanal)    | p-Tolualdehyde             |
| Butyraldehyde (Butanal) | Isovaleraldehyde         | Valeraldehyde (Pentanal)   |

CRM – Certified Reference Material

PT – Proficiency Testing

QR – QuiK Response

RM – Reference Material

**Q** All Waters ERA Air & Emissions PTs open quarterly. Quarterly months are January, April, July, and October.

# Metals

## Metals on Filter Paper

| CRM<br>Cat. #1125 | PT<br>Cat. #1025 | Q | QR<br>Cat. #1125QR |
|-------------------|------------------|---|--------------------|
|-------------------|------------------|---|--------------------|

One filter paper sample packaged in a 50 mm polystyrene petri dish containing a single 47 mm tissue quartz filter ready for use with EPA Method 29 or other applicable methods.

|                 |                  |
|-----------------|------------------|
| Antimony.....   | 25-250 µg/filter |
| Arsenic.....    | 20-250 µg/filter |
| Barium.....     | 20-250 µg/filter |
| Beryllium.....  | 10-250 µg/filter |
| Cadmium.....    | 10-250 µg/filter |
| Chromium.....   | 15-250 µg/filter |
| Cobalt.....     | 10-250 µg/filter |
| Copper.....     | 10-250 µg/filter |
| Lead.....       | 20-350 µg/filter |
| Manganese.....  | 10-250 µg/filter |
| Nickel.....     | 20-250 µg/filter |
| Phosphorus..... | 10-250 µg/filter |
| Selenium.....   | 20-250 µg/filter |
| Silver.....     | 30-250 µg/filter |
| Thallium.....   | 30-250 µg/filter |
| Zinc.....       | 20-250 µg/filter |

## Metals in Impinger Solution

| CRM<br>Cat. #1126 | PT<br>Cat. #1026 | Q | QR<br>Cat. #1126QR |
|-------------------|------------------|---|--------------------|
|-------------------|------------------|---|--------------------|

One impinger solution sample packaged in a 15 mL screw-top vial containing approximately 14 mL of standard concentrate for use with EPA Method 29, or other applicable methods.

|                 |               |
|-----------------|---------------|
| Antimony.....   | 0.25-20 µg/mL |
| Arsenic.....    | 0.2-20 µg/mL  |
| Barium.....     | 0.15-25 µg/mL |
| Beryllium.....  | 0.05-20 µg/mL |
| Cadmium.....    | 0.1-20 µg/mL  |
| Chromium.....   | 0.2-20 µg/mL  |
| Cobalt.....     | 0.1-25 µg/mL  |
| Copper.....     | 0.2-20 µg/mL  |
| Lead.....       | 0.2-20 µg/mL  |
| Manganese.....  | 0.1-20 µg/mL  |
| Nickel.....     | 0.15-30 µg/mL |
| Phosphorus..... | 0.15-25 µg/mL |
| Selenium.....   | 0.15-25 µg/mL |
| Silver.....     | 0.5-20 µg/mL  |
| Thallium.....   | 0.15-25 µg/mL |
| Zinc.....       | 0.15-25 µg/mL |

## Mercury on Filter Paper

| CRM<br>Cat. #1127 | PT<br>Cat. #1027 | Q | QR<br>Cat. #1127QR |
|-------------------|------------------|---|--------------------|
|-------------------|------------------|---|--------------------|

One 2 mL flame-sealed ampule containing approximately 2 mL of standard concentrate and a 50 mm polystyrene petri dish containing a single 47 mm glass fiber filter. Sample is ready for use with EPA Method 29, or other applicable methods.

|              |                |
|--------------|----------------|
| Mercury..... | 1-75 µg/filter |
|--------------|----------------|

## Mercury in Impinger Solution

| CRM<br>Cat. #1128 | PT<br>Cat. #1028 | Q | QR<br>Cat. #1128QR |
|-------------------|------------------|---|--------------------|
|-------------------|------------------|---|--------------------|

One impinger solution sample packaged in a 15 mL screw-top vial containing approximately 14 mL of standard concentrate for use with EPA Methods 29, 101a, or other applicable methods.

|              |               |
|--------------|---------------|
| Mercury..... | 0.9-200 ng/mL |
|--------------|---------------|

## Lead on Filter Paper

| CRM<br>Cat. #1129 | PT<br>Cat. #1029 | Q | QR<br>Cat. #1129QR |
|-------------------|------------------|---|--------------------|
|-------------------|------------------|---|--------------------|

One filter paper sample packaged in a 50 mm polystyrene petri dish containing a single 47 mm tissue quartz filter spiked with lead ready-for-use with EPA Method 12 or other applicable methods.

|           |                  |
|-----------|------------------|
| Lead..... | 20-350 µg/filter |
|-----------|------------------|

## Lead in Impinger Solution

| CRM<br>Cat. #1130 | PT<br>Cat. #1030 | Q | QR<br>Cat. #1130QR |
|-------------------|------------------|---|--------------------|
|-------------------|------------------|---|--------------------|

One impinger solution sample packaged in a 15 mL screw top vial containing approximately 14 mL of standard concentrate for use with EPA Method 12, or other applicable methods.

|           |               |
|-----------|---------------|
| Lead..... | 0.2-120 µg/mL |
|-----------|---------------|

## Chromium on Filter Paper

| CRM<br>Cat. #1131 | PT<br>Cat. #1031 | Q | QR<br>Cat. #1131QR |
|-------------------|------------------|---|--------------------|
|-------------------|------------------|---|--------------------|

One filter paper sample packaged in a 50 mm polystyrene petri dish containing a single 47 mm fiber film filter for use with CARB Method 425, or other applicable methods.

|                          |                |
|--------------------------|----------------|
| Total chromium.....      | 1-20 µg/filter |
| Hexavalent chromium..... | 1-20 µg/filter |

## Hexavalent Chromium in Impinger Solution

| CRM<br>Cat. #1132 | PT<br>Cat. #1032 | Q | QR<br>Cat. #1132QR |
|-------------------|------------------|---|--------------------|
|-------------------|------------------|---|--------------------|

One impinger solution sample packaged in a 15 mL screw top vial containing approximately 14 mL of standard concentrate for use with EPA Method 0061/7199, or other applicable methods.

|                          |             |
|--------------------------|-------------|
| Hexavalent chromium..... | 45-880 µg/L |
|--------------------------|-------------|

# Inorganics

## Hydrogen Halides & Halogens in Impinger Solution

|                   |                  |   |                    |
|-------------------|------------------|---|--------------------|
| CRM<br>Cat. #1140 | PT<br>Cat. #1040 | Q | QR<br>Cat. #1140QR |
|-------------------|------------------|---|--------------------|

Two impinger solution samples packaged in 15 mL screw-top vials containing approximately 14 mL of standard concentrate for use with EPA Methods 26, 26a, or other applicable methods.

|                        |              |
|------------------------|--------------|
| Total halides.....     | 15-1500 mg/L |
| Total halogens.....    | 10-200 mg/L  |
| Hydrogen chloride..... | 5-500 mg/L   |
| Hydrogen fluoride..... | 5-500 mg/L   |
| Hydrogen bromide.....  | 5-500 mg/L   |
| Bromine.....           | 5-100 mg/L   |
| Chlorine.....          | 5-100 mg/L   |

## Fluoride in Impinger Solution

|                   |                  |   |                    |
|-------------------|------------------|---|--------------------|
| CRM<br>Cat. #1141 | PT<br>Cat. #1041 | Q | QR<br>Cat. #1141QR |
|-------------------|------------------|---|--------------------|

One impinger solution sample packaged in a 15 mL screw-top vial containing approximately 14 mL of standard concentrate for use with EPA Methods 13a, 13b, 14, or other applicable methods.

|               |              |
|---------------|--------------|
| Fluoride..... | 1-50 mg/dscm |
|---------------|--------------|

## Nitrogen Oxide in Impinger Solution

|                   |                  |   |                    |
|-------------------|------------------|---|--------------------|
| CRM<br>Cat. #1142 | PT<br>Cat. #1042 | Q | QR<br>Cat. #1142QR |
|-------------------|------------------|---|--------------------|

One impinger solution sample packaged in a 15 mL screw-top vial containing approximately 14 mL of standard concentrate for use with EPA Method 7, or other applicable methods.

|                               |                  |
|-------------------------------|------------------|
| Oxides of nitrogen (NOx)..... | 100-2000 mg/dscm |
|-------------------------------|------------------|

## Sulfur Dioxide in Impinger Solution

|                   |                  |   |                    |
|-------------------|------------------|---|--------------------|
| CRM<br>Cat. #1143 | PT<br>Cat. #1043 | Q | QR<br>Cat. #1143QR |
|-------------------|------------------|---|--------------------|

One impinger solution sample packaged in a 15 mL screw-top vial containing approximately 14 mL of standard concentrate for use with EPA Method 6 and Method 8, or other applicable methods.

|                     |                 |
|---------------------|-----------------|
| Sulfur dioxide..... | 50-2000 mg/dscm |
|---------------------|-----------------|

## Sulfuric Acid & Sulfur Dioxide in Impinger Solution

|                   |                  |   |                    |
|-------------------|------------------|---|--------------------|
| CRM<br>Cat. #1144 | PT<br>Cat. #1044 | Q | QR<br>Cat. #1144QR |
|-------------------|------------------|---|--------------------|

One impinger solution sample packaged in a 15 mL screw top vial containing approximately 14 mL of standard concentrate for use with EPA Method 8, or other applicable methods.

|                    |               |
|--------------------|---------------|
| Sulfuric acid..... | 5-150 mg/dscm |
|--------------------|---------------|

## Ammonia in Impinger Solution

|                   |                  |   |                    |
|-------------------|------------------|---|--------------------|
| CRM<br>Cat. #1145 | PT<br>Cat. #1045 | Q | QR<br>Cat. #1145QR |
|-------------------|------------------|---|--------------------|

One impinger solution sample packaged in a 15 mL screw-top vial containing approximately 14 mL of standard concentrate for use with EPA CTM 027, or other applicable methods.

|               |             |
|---------------|-------------|
| Ammonium..... | 0.1-10 mg/L |
|---------------|-------------|

## Particulate Matter on Filter Paper

|                   |                  |   |                    |
|-------------------|------------------|---|--------------------|
| CRM<br>Cat. #1150 | PT<br>Cat. #1050 | Q | QR<br>Cat. #1150QR |
|-------------------|------------------|---|--------------------|

One filter paper sample packaged in a 50 mm polystyrene petri dish containing a single 47 mm tissue quartz filter ready for use with EPA Methods 5, 5A, 5B, 5D, 5F, or other applicable methods.

|                         |                  |
|-------------------------|------------------|
| Particulate matter..... | 50-600 mg/filter |
|-------------------------|------------------|

## Particulate Matter in Impinger Solution

|                   |                  |   |                    |
|-------------------|------------------|---|--------------------|
| CRM<br>Cat. #1151 | PT<br>Cat. #1051 | Q | QR<br>Cat. #1151QR |
|-------------------|------------------|---|--------------------|

One impinger solution sample packaged in a 250 mL polyethylene bottle containing approximately 250 mL of standard ready for use with EPA Methods 5, 5A, 5B, 5D, 5F, or other applicable methods.

|                         |              |
|-------------------------|--------------|
| Particulate matter..... | 140-675 mg/L |
|-------------------------|--------------|

CRM – Certified Reference Material

PT – Proficiency Testing

QR – QuiK Response

**Q** All Waters ERA Air & Emissions PTs open quarterly. Quarterly months are January, April, July, and October.



Learn more about Air & Emissions products