



10,000 uS/cm Conductivity -- 1 L

Item No. 01640

Conductivity calibration or verification standard
prepared from KCl and ultrapure water at
10000 uS/cm - 1L 6 month expiration



SPECIFICATIONS

Product Family	Process Water Consumables
Product Category	Conductivity
Item Type	CRM
Accreditations	ISO 17034
Lead Time (Business Days)	5
Storage	20 +/-10C (Ambient)
Shelf Life from date of manufacture (months)	6
Components per Pkg	1
Container Type	1L HDPE Bottle

Concentration	10,000 uS/cm
Volume	1 Liter
Weight	0.64 lbs
Application	Pharmaceutical
Preservative	Unpreserved