RANDOX

FRUCTOSAMINE CONTROL - LEVEL 3 (FRUC CONTROL 3)

CAT. NO. FR 2996 **LOT NO.** 596FR
SIZE: 3 x 1 ml **EXPIRY:** 2023-09-28
GTIN: 05055273203134

INTENDED USE
This product is intended for *in vitro* use in the quality control of the Randox Liquid Fructosamine assay on clinical chemistry systems.

SAFETY PRECAUTIONS AND WARNINGS
Human source material, from which this product is derived, has been tested at donor level for the Human Immunodeficiency Virus (HIV1 & HIV2) antibody, Hepatitis B surface antigen (HBsAg) and the Hepatitis C Virus (HCV) antibody and found to be **NON-REACTIVE**.
However, since no method can offer complete assurance as to the absence of infectious agents, this material and all patient samples should be handled as though capable of transmitting disease.
For *in vitro* diagnostic use only.

STORAGE AND STABILITY
Unopened fructosamine control is stable until the expiry date printed on the product label, when stored between +2°C and +8°C.
Once reconstituted the components of the serum are stable for 28 days at +2°C to +8°C, and 1 month at –20°C when frozen once.

PREPARATION FOR USE
Open the vial carefully, avoiding any loss of the material and reconstitute with 1 ml of distilled water. Replace the rubber stopper, close the vial and leave to stand for 30 minutes before use. Ensure that all traces of dry material are dissolved by swirling gently.

MATERIALS PROVIDED
Fructosamine Control - Level 3 (3 x 1 ml)

MATERIALS REQUIRED BUT NOT PROVIDED
Distilled water
Volumetric pipette

VALUE ASSIGNMENT
Due to the variation caused by test equipment, test reagents and laboratory technique, the quoted ranges are provided for guidance. It is recommended that these ranges are used until each laboratory has established its own ranges, based on individual laboratory requirements.

The control value was assigned relative to human serum glycated with ^{14}C-glucose and is the mean of at least 30 replicate determinations, using various clinical chemistry systems in a single laboratory.

RANDOX ENZYMATIC METHOD

	TARGET	RANGE
FRUCTOSAMINE	728 µmol/l	582 – 874 µmol/l

SIEMENS ENZYMATIC FOR ADVIA® AND ATELLICA CH®

	TARGET	RANGE
FRUCTOSAMINE	866 µmol/l	693 – 1039 µmol/l

EC REP Randox Teoranta, Meenmore,
Dungloe, Donegal,
F94 TV06, Ireland

14 Oct '22 me

THIS PAGE IS INTENTIONALLY BLANK

Randox Laboratories Ltd, 55 Diamond Road, Crumlin, County Antrim, BT29 4QY, United Kingdom
T +44 (0) 28 9442 2413 **F** +44 (0) 28 9445 2912 **E** applications@randox.com **I** www.randox.com