

# **Mobile Phase 1 Reagent**

REF	01-03-0042	Fill 940mL
REF	01-03-0040	Fill 3.8L

Pour d'autres langues
Für andere Sprachen
Para otras lenguas
Per le altre lingue
Dla innych języków

Para outras línguas	
Για τις άλλες λώσσες	
För andra språk	
For andre språk	

www.trinitybiotech.com

# INTENDED USE

Mobile Phase 1 Reagent is designed for use with the Trinity Biotech ultra2 systems for the separation and quantitation of normal and variant hemoglobin species in blood. For in vitro diagnostic use only. IVD

# SUMMARY AND EXPLANATION OF TEST

Hemoglobin Variants - Assessment of hemoglobin variants has proven useful in the identification of altered hemoglobin commonly associated with disorders called hemoglobinopathies.

Reagents are performance validated to assure accuracy and precision with the Trinity Biotech assay and system for the screening for hemoglobin variants.

Reagent is ready for use.

# STORAGE AND STABILITY

4 Store at ambient temperatures (2 – 28°C). Do not allow to freeze.

The reagent is stable until the expiration date indicated on the label when kept tightly closed and protected from extreme environmental conditions.



See the container label for the expiration date. DO NOT USE after the expiration date.

#### PRECAUTIONS

For in vitro diagnostic use only. Avoid skin contact. Consult the product SDS for safety information. This reagent is used in conjunction with blood testing equipment and warrants handling under universal precaution procedures for safety.

ORDERING INFORMATION				
Reference No.	Item	Quantity		
01-03-0042	Mobile Phase 1	940mL		
01-03-0040	Mobile Phase 1	3.8L		





2



1

Consult Instructions for Use

European Conformity

CE





For in vitro Diagnostic Use

Temperature limitation



Caution, consult accompanying documents



Trinity Biotech Kansas City, MO 64132 USA Tel. 1 800-325-3424 Fax: 1 816-361-1974

Trinity Biotech plc Bray Co. Wicklow, Ireland Tel. 353 1 2769800 EC REP

Fax 353 1 2769888 www.trinitybiotech.com